This item was submitted to [Loughborough's Research Repository](#) by the author.
Items in Figshare are protected by copyright, with all rights reserved, unless otherwise indicated.

Development of a UHPLC-MS/MS (SRM) method for the quantitation of endogenous glucagon and dosed GLP-1 from human plasma

PLEASE CITE THE PUBLISHED VERSION

<http://dx.doi.org/10.4155/bio-2017-0021>

PUBLISHER

© 2017 Future Science Ltd

VERSION

AM (Accepted Manuscript)

PUBLISHER STATEMENT

This work is made available according to the conditions of the Creative Commons Attribution-NonCommercial-NoDerivatives 4.0 International (CC BY-NC-ND 4.0) licence. Full details of this licence are available at:
<https://creativecommons.org/licenses/by-nc-nd/4.0/>

LICENCE

CC BY-NC-ND 4.0

REPOSITORY RECORD

Howard, James W., Richard G. Kay, Ben Jones, Jaimini Cegla, Tricia Tan, Steve Bloom, and Colin Creaser. 2019. "Development of a UHPLC-MS/MS (SRM) Method for the Quantitation of Endogenous Glucagon and Dosed GLP-1 from Human Plasma". figshare. <https://hdl.handle.net/2134/26217>.

Development of a UHPLC-MS/MS (SRM) method for the quantitation of endogenous glucagon and dosed GLP-1 from human plasma

James W Howard^{1,2†}, Richard G Kay³, Ben Jones⁴, Jaimini Cegla⁴, Tricia Tan⁴, Steve Bloom⁴ and Colin S Creaser²

1 LGC Limited, Newmarket Road, Fordham, Cambridgeshire,
CB7 5WW, UK

2 Centre for Analytical Science, Department of Chemistry, Loughborough University,
Leicestershire, LE11 3TU, UK

3 University of Cambridge, Institute of Metabolic Science, Cambridgeshire CB2 0QQ, UK

4 Imperial College, Department of Investigative Medicine, Hammersmith Hospital Campus,
Du Cane Road, London, W12 0NN, UK

† Author for correspondence. Tel: +44 (0) 1638 720 500. Fax: +44 (0)1638 724 200

Email: james.howard@lgcgroup.com

Abstract

Background: The performance of glucagon and GLP-1 immunoassays is often poor, but few sensitive LC-MS/MS methods exist as alternatives. **Results:** We established the first multiplexed LC-MS/MS method avoiding immunoenrichment for the quantitation of endogenous glucagon (LLOQ 15 pg/mL) and dosed GLP-1 (LLOQ 25 pg/mL) in human plasma. Specificity of endogenous glucagon quantitation was assured using a novel approach with a supercharging mobile phase additive to access a sensitive qualifier SRM. Endogenous glucagon concentrations were within the expected range, and showed good reproducibility after extended sample storage. A cross-validation against established immunoassays using physiological study samples demonstrated some similarities between methods. **Conclusion:** The LC-MS/MS method offers a viable alternative to immunoassays for quantitation of endogenous glucagon, dosed glucagon and/or dosed GLP-1.

Keywords: glucagon, GLP-1, endogenous, plasma, LC-MS/MS, immunoassay, cross-validation, supercharging mobile phase additive, *m*-NBA

Introduction

Glucagon and glucagon-like peptide-1 (GLP-1) are peptide hormones encoded by the proglucagon gene, and are released from the gene product via tissue specific post translational processing (Figure 1).

Gut/ Brain	Glicentin					GLP-1 (7-36)		GLP-2	
		Oxyntomodulin							
Position	1-32	33-61	62-69	70-71	72-77	78-107	108-125	126- 158	159-160
Amino acid	RSLQDTEEKSRS FSASQADPLSDP DQMNEDKR	HSQGTFTS DYSKYLDS RRAQDFV QWLMNT	KR NR NN IA	KR	HD EF ER	HAEGTFTSDV SSYLEGQAAKE FIAWLVKGR	GRRDFPE EVAIVEEL GRR	HADGSFSDEM NTILDNLAAARD FINWLIQTKITD	RK
Pancreas		Glucagon			Major pro-glucagon fragment				

Figure 1 Proglucagon sequence and major processing products in the in the gut/brain and pancreas

Glucagon is released from pancreatic α cells and is a counter regulatory hormone that responds to hypoglycemia and fasting by stimulating glycogenolysis and gluconeogenesis, as well as hepatic fatty acid β -oxidation and ketogenesis [1]. It is also a biomarker for diseases such as diabetes and neuroendocrine tumours [2]. GLP-1 is released *via* secretion from intestinal L cells, and has primary roles in enhancing the β -cell insulin response to eating, enhancing β -cell survival, inhibiting gastric emptying, inhibiting glucagon secretion, and suppressing appetite [1][3][4], as well as being of interest as a biomarker [5]. Pharmacological administrations of glucagon are known to increase energy expenditure [1], and therefore it is of interest along with GLP-1 for the development of obesity treatments [6][7].

To study the physiological role of glucagon and GLP-1, and exploit their use as biomarkers, precise and accurate methods for determining their plasma concentrations are required. These are also needed to determine their pharmacokinetics in studies where these are dosed [6][7]. Traditionally such peptides are quantified using immunoassays. However precision and accuracy can be poor, as can be the correlation between assays [8][9][10][11][12][13]. This is often attributed to the potential for antibodies to cross-react with similar compounds, including inactive degradation fragments and metabolites. Specificity is particularly challenging for GLP-1 assays due to the large number of isoforms present that may cross-react. GLP-1 1-37, 7-37, and 9-37 are produced from differential cleavage of the pro-glucagon precursor [11]. In humans these primarily exist in C-terminal amidated isoforms (GLP 1-36NH₂, 7-36NH₂, and 9-36NH₂) [4] [11][14]. GLP-1 7-36NH₂ is the biologically active form and is referred to by the unqualified GLP-1 nomenclature. Similarly, some glucagon

immunoassays assays are known to cross-react with proglucagon products glicentin and oyntomodulin [9]. Glucagon metabolites, some of which have only been recently reported [15], may also cross-react.

These concerns cast doubt on the integrity of some of the data in the literature. Cross-validations to help assess the performance of different kits and laboratories can be impractical, as kits can be expensive, resources may not be available for training, and assay specific equipment may be required. For example only one of seven glucagon immunoassays recently evaluated used a standard microplate reader [9]. In addition radioimmunoassays (RIA) necessitate additional health and safety precautions during set-up and require specialised disposal of radioisotopes.

LC-MS/MS based methodologies can help to overcome some of these challenges. For example, specificity can be improved by monitoring SRM transitions that incorporate the full length peptide and a related fragment ion, which will exclude many structurally similar compounds, and extraction methodologies and chromatographic separations can be tailored to the peptides of interest. Furthermore qualifier SRM transitions can be monitored, provided sufficiently sensitive transitions can be determined, to ensure results are consistent with those achieved from the quantitation SRM [16][17][18][19]. Inter-lab cross validations are also easier as methods can be transferred between LC-MS/MS systems with limited re-optimisation.

However the application of LC-MS/MS for glucagon and GLP-1 quantitation has been limited, primarily due to sensitivity challenges. Typical human plasma glucagon concentrations are in the region of 30 pg/mL, which can increase by 30 - 100 pg/mL as a result of hypoglycemia, but decrease to ~3 – 7 pg/mL during hyperglycaemia [9]. GLP-1 is typically present at lower plasma concentrations, at approx. 13 pg/mL for fed and 2 pg/mL for fasted subjects [20].

Despite these challenges, some sensitive LC-MS/MS methods have been reported. We described a method capable of routinely quantifying glucagon to 25 pg/mL using 400 µL plasma [21]. Another group recently described a multiplexed LC-MS/MS method capable of quantifying endogenous glucagon to 2.7 pg/mL and GLP-1 to 2.6 pg/mL using 500 µL plasma [20], which was a refinement of their previous method using 1000 µL [22]. However these methods used immunoaffinity enrichment with magnetic beads coated with anti-analyte antibodies, adding expense and complexity to the method, and the methods had long LC run times (ca 15 minutes) due to the use of microflow LC to obtain the required

sensitivity. Methods for GLP-1 quantitation avoiding immunochemistry have much higher LLOQs, for example 66 pg/mL in a recent paper [23].

In this study we refined and expanded our previous method, to improve sensitivity, robustness and throughput, and to add GLP-1 as a secondary analyte. The method was qualified using fit-for-purpose criteria based on key experiments from FDA [24] and EMA [25] bioanalytical validation guidelines. We also describe the novel use of a supercharging mobile phase additive, meta nitrobenzyl alcohol (*m*-NBA), to obtain a sensitive qualifier SRM transition to ensure specificity of endogenous glucagon quantitation. The method was cross-validated against two established immunoassays for each analyte using the same physiological study sample set.

Experimental

Chemicals and Materials

Certified human glucagon (HSQGTFTSDYSKYLDSRRAQDFVQWLMNT) was obtained from EDQM (Strasbourg, France). The analogue internal standard (IS) (des-thr⁷-glucagon) (HSQGTFSKYLDSRRAQDFVQWLMNT) and stable isotope labelled (SIL) internal standard (HSQGT-[¹³C₉,¹⁵N]F-TSDYSKYLDSRRAQDFVQW-[¹³C₆,¹⁵N]L-MNT) were obtained from Bachem (Bubendorf, Switzerland). GLP-1 (7-36) amide (HAEGTFTSDVSSYLEGQAAKEFIAWLVKGR - NH₂) was also obtained from Bachem.

Water was produced by a Triple Red water purifier (Buckinghamshire, U.K.). All chemicals and solvents were HPLC or analytical reagent grade and purchased from either Fischer Scientific (Loughborough, UK) or Sigma Aldrich (St Louis, MO USA).

Preparation of Stock and Spiking Solutions

Stock solutions of glucagon and glucagon internal standards (1 mg/mL), and stocks of GLP-1 (0.1 mg/mL), were prepared in borosilicate glass vials using surrogate matrix [MeOH: H₂O: Formic acid (FA): Bovine serum albumin (BSA), (20/80/0.1/0.1, v/v/v/w); typically 200mL H₂O, 800mL MeOH, 1mL FA and 1 g BSA]. Combined glucagon and GLP-1 working solutions were prepared by dilution with this solvent to create nine calibration standard spiking solutions (300, 500, 900, 2000, 6000, 12000, 20000, 37000 40000 pg/mL), and six quality control spiking solutions (300, 500, 900, 1500, 4000, 35000 pg/mL). A mixed internal standard working solution (ISWS) containing both analog glucagon and SIL glucagon was similarly prepared at 20 ng/mL. The stock and working solutions were prepared to a volume of 10 mL and were stored at -20 °C when not in use, as we previously demonstrated that glucagon solutions were stable under these conditions [21].

Method Development

The previously reported method [21] was refined to increase sensitivity, robustness and throughput. Various changes to the extraction procedure and LC conditions were investigated, as summarised in Supplementary information Table 1. SRM transitions corresponding to the SIL glucagon IS and GLP-1 were identified and optimised. Mobile phases were modified with various proportions of the supercharging additive *m*-NBA (meta nitrobenzyl alcohol) and formic acid to attempt to improve sensitivity. Further information is presented in Supplementary Information Sections 1-5. The resulting extraction and LC-MS/MS methods used for the qualification are detailed below.

Quantitation Strategy

A surrogate matrix based approach was used for glucagon quantitation. Plasma QCs, to represent samples, were created by diluting spiking solutions 20-fold into EDTA plasma to 200 (MED) and 1750 (HIGH) pg/mL, and the QC concentration adjusted for the endogenous concentration. These were extracted and analysed according to the qualified extraction and LC-MS/MS methods described below. Calibration standards, QC LLOQs (precision and accuracy batches only) and QC LOWs were then prepared by spiking 20 µL of the appropriate spiking solution into the collection plate, along with 20 µL of ISWS and 160 µL surrogate matrix. Taking into account the 2-fold concentration experienced by plasma samples (400 µL of plasma sample is reconstituted into 200 µL of solvent) this gave final calibration levels of 15, 25, 45, 100, 300, 600, 1000, 1850, and 2000 pg/mL, and final QC levels of 15 (QC LLOQ) and 45 (QC LOW) pg/mL.

Endogenous GLP-1 could not be detected, and therefore GLP-1 quantitation was performed as per an exogenous compound with calibrants and QCs prepared in the sample matrix (EDTA Plasma). Spiking solutions were diluted 20-fold into plasma to produce calibrants at 25, 45, 100, 300, 600, 1000, 1850 and 2000 pg/mL and QC concentrations at 25 (LLOQ), 75 (LOW), 200 (MED), and 1750 (HIGH) pg/mL. These were extracted according to the qualified extraction method.

Qualified Extraction Method

Plasma sample (EDTA) (400 µL) was placed into a 2mL 96 well plate and 20 µL of ISWS was added to all non-blank samples. This step was performed on ice to ensure analyte stability [26]. The plate was vortex mixed, and samples precipitated using 1.1 mL of ACN:H₂O:NH₃ (72:25:0.1, v/v/v), vortex mixed, sonicated for 5 minutes, and then centrifuged for 10 minutes at 2300 x g. 1.2mL of supernatant was transferred using an automated liquid handling system (Quadra Tower, TomTec, Connecticut, USA) to a 2 mL plate and evaporated to dryness at 40°C under nitrogen (ca 90 minutes). Samples were reconstituted in 800 µL 2% NH₃ (aq) and then vortex mixed, before being extracted using solid phase extraction (SPE) as per our previously reported method [21] A Bond Elut Plexa 96 round-well SPE plate (30 mg) was conditioned using 1 mL MeOH, then equilibrated with 1 mL H₂O. The samples were loaded, washed with 1 mL 5% MeOH (aq), eluted with 2 x 225 µL ACN:H₂O:FA (75:25:0.1, v/v/v) into a 1 mL Lo-bind plate, and then evaporated under nitrogen at 40°C, before being reconstituted in 200 µL 0.2% FA (aq). The plate was centrifuged for 10 minutes at 2300 x g, and 40 µL of sample injected on to the LC-MS/MS system for analysis, or the plate stored at 4°C awaiting injection. The entire process took 1 working day.

Qualified LC-MS/MS method (formic acid phases)

The LC-MS/MS system consisted of a Waters Acquity UPLC system (Waters Corporation, Massachusetts, USA) coupled to an AB SCIEX 5500 (Applied Biosystems / MDS SCIEX, Ontario, Canada) with an electrospray ion source.

Glucagon was separated on a Waters Acquity UPLC BEH C18 1.7 μm (2.1 x 100 mm) column maintained at 60 °C. The mobile phase consisted of (A) 0.2% FA in acetonitrile (ACN) and (B) 0.2% FA (aq). The gradient for glucagon elution was 22 - 32% A over 2 minutes, as in our previous method [21], followed by a gradient of 32 - 38% over 0.8 minutes for GLP-1 elution. The column was then cleaned with 95% A for 0.75 minutes then re-equilibrated at the starting conditions for 0.05 minutes. The flow rate was 0.8 mL/min and the total gradient time was 3.6 minutes per sample.

The mass spectrometer was operated in positive ion mode with an electrospray voltage of 5500 V. The source temperature was 600°C, the curtain, N₂ collision, GS1, and GS2 gases were set to 40, 8, 60, 40 psi respectively. The Q1 and Q3 quadrupoles were both operated at unit resolution. Entrance potentials of 10 V and collision exit cell potentials of 13 V were used. MS periods were used to maximise dwell times, and therefore sensitivity. The selected reaction monitoring (SRM) transitions 697.5→694.0, 677.4→673.8 and 702.2→698.8 for glucagon, analogue glucagon IS, and SIL glucagon IS were monitored between 0 – 2 minutes. These transitions were attributed to $[\text{M}+5\text{H}^+]^{5+} \rightarrow [\text{M}+5\text{H}^+-\text{NH}_3]^{5+}$ in each case. Transitions used the optimal CE of 20, 18, and 20 respectively and all used dwell times of 60 ms with declustering potential (DP) of 80. The SRM transitions 660.6→656.9 and 660.6→752.0, which were attributed to $[\text{M}+5\text{H}^+]^{5+} \rightarrow [\text{M}+5\text{H}^+-\text{NH}_3]^{5+}$ and $[\text{M}+5\text{H}^+]^{5+} \rightarrow \text{y}_{20}^{3+}$ respectively, were monitored for GLP-1 between 2 - 3.6 minutes. These transitions used the optimal CE of 16 and 22, and DP of 100 and 110 respectively and all used dwell times of 100 ms.

Qualified LC-MS/MS Method (m-NBA mobile phases)

The same LC-MS/MS systems were used as above. However the mobile phase consisted of (A) 0.01% (FA) in ACN with 0.05% *m*-NBA and (B) 0.01% FA (95/5 H₂O/ACN) with 0.05% *m*-NBA. The addition of ACN to the aqueous phase enabled solubility of the *m*-NBA. The gradient for glucagon elution was 18–28% A over 2 minutes, then 28 -34% over for 0.8 minutes for GLP-1 elution, before being cleaned and re-equilibrated as above. MS settings were as above, unless stated differently below. The selected reaction monitoring (SRM) transitions 581.5→578.5, 564.7→561.8 and 585.5→582.4 for glucagon, analogue glucagon IS and SIL glucagon IS were monitored between 0 – 1.75 minutes. These transitions were

attributed to $[M+6H]^+ \rightarrow [M+6H-NH_3]^+$ in each case. Transitions used the optimal CE of 13, 12, 13 respectively and all used DP of 80 and dwell times of 60 ms. The SRM transitions 550.6→601.6 and 550.6→639.3 were monitored for GLP-1 between 1.75 - 3.6 minutes. These transitions were attributed to $[M+6H]^+ \rightarrow y_{16}^{3+}$ and $[M+6H]^+ \rightarrow y_{17}^{3+}$ respectively. Transitions used the optimal DP of 80 and 70 respectively and both used CE of 19 and dwell times of 100 ms.

Qualification Procedure

The qualification experiments were based on key experiments described in the FDA [24] and EMA [25] bioanalytical guidelines. Analyst v 1.6.2 (Applied Biosystems / MDS SCIEX, Ontario, Canada) was used to process data and to construct calibration lines. Duplicate calibration standards were analysed within each batch, and calibration lines were constructed using peak area ratio-concentration plots with linear regression and $1/x^2$ weighting.

For glucagon our acceptance criteria required that the accuracy of $\geq 75\%$ of standards in each batch was within 15% (20% at the LLOQ) of the nominal concentration, with those outside this excluded from the calibration line. For precision and accuracy batches QCs needed mean accuracy within 20%, with precision (CV) $\leq 20\%$ at each level. In other batches $\geq 2/3$ of the individual QCs needed accuracy within 20%, with at least one QC at each level. For GLP-1 similar acceptance criteria was used, except that a 25% criteria was used for precision and accuracy at all levels.

Precision and accuracy of the method was assessed by analysis of replicate (n=6) QC samples at at least four different concentrations within a batch. Selectivity was determined by inspecting chromatograms from six independent plasma EDTA samples for the presence of potentially interfering peaks.

The effect of the presence of matrix to the response of the analyte and internal standard was determined from six independent plasma EDTA samples. These were extracted and post spiked at the 200 pg/mL level, and compared to the mean response from samples in surrogate matrix taking into account endogenous concentrations. In addition QCs were fortified at 200 pg/mL in six independent Plasma EDTA matrices (1 aliquot of each), and in lithium heparin plasma (n=6).

Stability in matrix was determined by spiking a plasma sample at 200 pg/mL and incubating it at room temperature or ice for 4 hr 50 minutes before extraction (n=6 replicates), and compared to a sample immediately extracted. The ability to re-inject extracts was assessed

at the P&A QC levels after 42 days storage at 4°C. Analyte recovery was evaluated by comparing samples (n=6) spiked with 200 pg/mL analyte, with extracts post spiked at this level to represent 100% recovery. The 10-fold dilution of a QC HIGH sample (1750 pg/mL)(n=6) with EDTA plasma was used to demonstrate whether low samples volumes could be analysed.

Collection of samples from volunteers to assess endogenous glucagon concentrations

Blood was collected from 11 male and 11 female healthy volunteers using glass plasma collection tubes (5 mL, EDTA anticoagulant) , obtained from BD (Oxford, UK). Tubes were placed on ice immediately after sample collection and then centrifuged at 2300 x g for 10 minutes to obtain plasma, which was stored at -80°C. Plasma was collected at the start of the working day and volunteers were not asked to change their usual eating regime.

Glucagon concentrations were determined using the qualified method with formic acid phases. Concentrations were determined upon the first freeze-thaw of the aliquot, as multiple cycles have been shown to affect quantitation [21]. The ability to extrapolate below the 15 pg/mL for glucagon limit of quantitation was assessed by determining precision and accuracy of surrogate matrix QCs (n=6) diluted to 10 and 7.5 pg/mL levels.

Samples containing endogenous levels of glucagon above the lower limit of quantitation (15 pg/mL) were reanalysed after storage for up to 249 days at -80°C to assess incurred sample reproducibility (ISR). Similarly a plasma pool from one male and one female individual was created from stored samples and analysed immediately after pooling and after additional storage for 42 days at -80°C.

Finally, extracts from a selection of samples (n=9), and corresponding calibration standards and QC samples were reanalysed with the qualified method using *m*-NBA mobile phases to access a sensitive qualifier SRM to assure the methods selectivity.

Collection of physiological study samples

A 20 mL blood sample was collected from ten fasting healthy volunteers (West London National Research Ethics Committee, ref. 11/LO/1782), one volunteer dosed with glucagon (15 pmol/kg/min iv) (London Central Ethics and Research Committee (13/LO/0925), and two volunteers dosed with GLP-1 (16 pmol/kg/min sc) (13/LO/1510). Immediately after collection, 2 ml samples were decanted into various collection tube types: 1) K3 EDTA anticoagulant (ref 368860 – Becton Dickinson Vacutainer System, NJ, USA) with the addition of aprotinin (Bayer, Newbury, UK; final concentration 250 kallikrein inhibitor units/ml) and a DPPIV inhibitor (Ile-Pro-Ile, Sigma-Aldrich, Dorset, UK; final concentration 10 µg/ml); 2) K3 EDTA anticoagulant with the addition of aprotinin; 3) K3 EDTA anticoagulant with the addition of a

DPPIV inhibitor; 4) K3 EDTA anticoagulant alone; 5) lithium heparin (ref 367883 – Becton Dickinson Vacutainer System, NJ, USA) with the addition of aprotinin and a DPPIV inhibitor; 6) lithium heparin with the addition of aprotinin; 7) lithium heparin with the addition of a DPPIV inhibitor and 8) lithium heparin alone. Following collection, tubes were placed on ice and centrifuged at 2300 x g for 10 minutes to obtain plasma, which was stored at -80°C. Two further aliquots, collected in K3 EDTA tubes with aprotinin and DPPIV inhibitor, underwent 1 and 2 further freeze/thaw cycles, respectively.

Analysis of physiological study samples

Samples of the same stabiliser type and anticoagulant from three of the ten fasting healthy volunteers were pooled together. One aliquot was reserved for endogenous level determination and a second aliquot was spiked with 200 pg/mL analyte. Samples were analysed using the qualified LC-MS/MS method with formic acid mobile phases to assess the effect that a change of sample matrix had upon quantitation.

The majority of the samples from five of the remaining seven fasting healthy volunteers (endogenous glucagon samples) were analysed by the LC-MS/MS method and by two immunoassays (50/50, 50/50 and 45/50 for the LC-MS/MS, first, and second immunoassay respectively). The first immunoassay was a Homogenous Time Resolved Fluorescence (HTRF®) sandwich immunoassay (Cisbio, Codolet, France) (analytical sensitivity = 12.0 pg/mL). The second was a Milliplex MAP (multi-analyte profiling) Human Metabolic Hormone Magnetic Bead Panel (Merck Millipore, Darmstadt, Germany) (minimum detectable concentration (MinDC) = 13.0 pg/mL), which involved a sandwich immunoassay with fluorescence detection. Samples from the other two fasting healthy volunteers were only analysed using one method format, and are therefore not considered further. The same formats were also used to analyse samples from the volunteer dosed with glucagon. Samples from the two volunteers dosed with GLP-1 were similarly analysed using the LC-MS/MS method, and by either a Millipore active GLP-1 ELISA (Lowest standard = 6.6 pg/mL) or the Milliplex MAP (MinDC = 1.2 pg/mL) described above.

The various measures of immunoassay sensitivity quoted by the manufactures (i.e. analytical sensitivity/ MinDC/ lowest calibration point) were taken as equivalent to the lower limit of quantitation (LLOQ) measurement of the LC-MS/MS method for the purpose of this study. For all immunoassays, intra-assay CVs were <10% across the working range.

Results and Discussion

Method Development –Re-optimisation

Initial development focused on re-optimisation of the glucagon method described previously [21] to attempt to improve sensitivity, robustness, and throughput for glucagon analysis. GLP-1 was the secondary analyte for the assay, and was not introduced at this stage. A summary of the re-optimisation is presented in Supplementary Information Table 1, and further details are given in Supplementary Information Sections 1-5.

The protein precipitation solvent volume used to extract the 400 µL plasma samples was lowered from 3.2 mL to 1.1 mL. Such low plasma: solvent ratios were not previously investigated as they would usually lead to high matrix effects, however the subsequent use of SPE as part of the 2D extraction procedure minimised these. This alteration allowed the assay to be transferred from 5 mL tubes into 96 well plates, enabling the use of an automated liquid handling system to transfer the protein precipitation supernatant further reducing extraction time and increasing robustness. Additionally it reduced the supernatant evaporation time to ca 90 minutes, halving the extraction time to 1 working day, which is similar to the incubation period for some glucagon immunoassays [9].

Following SPE the reconstituted samples contained less insoluble material than observed previously [21], and upon analysis an interference peak (1.6 min) close to glucagon's retention time (1.7) was eliminated (Supplementary Information Figure 2a vs b). This may be due to the lower transfer of solid protein precipitate using the automated system than with a manual transfer. The reduced matrix content associated with the modified method also eliminated the need for a lengthy column clean, shortening the LC gradient time to 3.6 minutes and increasing analyte signal in the plasma extracts.

A SIL glucagon IS was acquired to investigate whether this improved performance over the analogue IS used previously. Performance was similar under standard conditions, however the SIL IS gave better performance when matrix effects were high, and it was therefore selected to maximise robustness. The other parameters investigated did not result in an improvement of the method.

Method Development –Addition of GLP-1

LC-MS analysis of GLP-1 showed that the $[M+3H]^{3+}$, $[M+4H]^{4+}$ and $[M+5H]^{5+}$ charge states (m/z = 1009.5, 825.1, and 660.0) were observed using formic acid based mobile phases. The latter two being most intense, and of similar intensity to each other (Figure 2a). As with glucagon [21], very little fragmentation was observed at lower collision energies and once a critical energy was reached numerous low intensity products were observed. Following optimisation and evaluation of sensitivity and linearity, the 660.6→656.9 and 660.6→752.0 SRM transitions were found to be similarly optimal, with sufficient sensitivity obtained at 25 pg/mL for extracted plasma samples. Previously, either the 825.4 → 946.3 transition alone [20][22], or the 825.2→946.4 transition summed with 825.2 → 643.2 [23] were found optimal. Differences in the LC-MS/MS systems used, including the use of nano or microflow LC in the previous studies, as well as variations in the nature of samples used for evaluation, may help to explain the discrepancies.

Method Development –Supercharging Mobile Phase Additives

One of the challenges for sensitive LC-MS/MS peptide bioanalysis is that signal is split over multiple charge states. Supercharging mobile phase additives, for example *m*-NBA, are known to alter the charge state distribution favouring higher charge states [27]. This is thought to be due to an increase in surface tension of the droplets in the electrospray source, decreasing the radius prior to columbic explosion, and consequently leading to charge concentration [28]. Sensitivity increases may result if the charge state distribution is reduced, or if the signal is increased due to improved analyte ionisability, as we have demonstrated for one large peptide in plasma extract [27]. However for glucagon whilst we previously demonstrated alterations in the charge state distribution (3, 4, 5+ to 4, 5, 6+) and an improvement in analyte signal in extracts, no overall increase in sensitivity (signal to noise) resulted due to similar increases in background noise [27].

It was decided to re-investigate the use of *m*-NBA to improve glucagon sensitivity with the re-optimised extraction method and to more fully investigate the effect of altering the *m*-NBA and FA content of the mobile phases. Furthermore it was planned to use the *m*-NBA method to provide a sensitive qualifier SRM to ensure assay specificity as described later, so an optimal method was required. The effect on GLP-1 was investigated, which similarly to glucagon demonstrated a shift in charge state distribution (Figure 2b).

Figure 2- GLP-1 Full Scan MS Spectra- a) Formic acid mobile phases (Mobile A= 0.2% FA (ACN) Mobile B= 0.2% FA (aq)), b) Supercharging (*m*-NBA) mobile phases (Mobile A= ACN: *m*-NBA:FA (100:0.1:0.1), Mobile B= H₂O:ACN:FA:*m*-NBA (95:5:0.1:0.1))

The charge state distribution and signal intensity of glucagon and GLP-1 were affected by changes in the *m*-NBA and FA composition of the mobile phases in both solution and plasma extracts (Supplementary Information 5.1 – 5.3). Initial results suggested that higher levels of *m*-NBA favoured higher charge states (Supplementary Information 5.1.1) with the 5+ and 6+ charge state of glucagon being most dominant at 0.05% and 0.10% *m*-NBA respectively. However further experiments suggested that 0.05% *m*-NBA was optimal for both charge states (Supplementary Information 5.2). Significant changes in signal intensity

occurred with the alteration of mobile phase formic acid content, which did not occur when using standard mobile phases (Supplementary Information 5.1.2 vs 3.2). Interestingly mobile phases modified with low amounts of formic acid (e.g. 0.01%) gave optimal signal for both analytes, regardless of whether they were in solution or plasma matrix, although the absence of acid led to very low signal demonstrating its necessity for efficient ionisation. Overall mobile phases modified with 0.05% *m*-NBA 0.01% FA gave optimal signal in plasma and extract for both analytes.

Plasma extracts were analysed using both the optimal *m*-NBA and standard formic acid mobile phase methods (Supplementary Information 5.3). Similar sensitivity (signal-to-noise) was obtained using both conditions for glucagon and GLP-1, suggesting that *m*-NBA the method could provide a sensitive qualifier SRM. Method performance using *m*-NBA modified phases were further characterised during the establishment.

Quantitation Strategy

The endogenous nature of glucagon presents additional challenges as an authentic matrix free of analyte cannot be acquired, and therefore approaches such as standard addition, surrogate matrix or surrogate analyte must be considered as described previously [21]. This assay used the surrogate matrix based approach for glucagon quantitation, where calibration standards are prepared in an analyte free surrogate matrix, allowing low levels of glucagon in plasma samples to be quantified without extrapolation. Ideally surrogate calibrants and surrogate QC samples would be extracted alongside plasma samples, but this was not possible due to non-specific binding of the analyte in surrogate matrix to the extraction materials. Therefore non-extracted surrogate samples were used, which were added to the plate after plasma samples had been extracted.

Similarly, in such strategies it is common to evaluate the performance of plasma samples diluted with surrogate matrix to demonstrate parallelism between surrogate and authentic matrix [25][29][30]. However, this was not possible as dilution prevented the protein precipitation, which was required in the extraction procedure. Therefore the performance of QCs spiked with analyte on top of the endogenous levels against the surrogate calibration line was considered sufficient.

Acceptable sensitivity for is usually demonstrated by assessing whether the analyte response at the LLOQ level is at least 5 times the average response due to background noise [25]. It is then assumed that an unknown sample at the LLOQ concentration would also have a similarly acceptable response. However, this will not necessarily be the case for surrogate matrix assays, due to differences in the recovery and matrix suppression between

the surrogate and authentic matrices. The previous method [21] had a mean analyte recovery of 51.2% and matrix suppression (matrix factor=0.746), and therefore the signal-to-noise (S/N) required at the LLOQ was 13.1 to ensure that S/N for an authentic sample at the LLOQ level ≥ 5 (assuming an unchanged background level). However using the re-optimised extraction method the IS response from surrogate matrix and plasma samples were similar (Supplementary Information 5.4) demonstrating that recovery losses during plasma extraction were compensated by matrix enhancement. The increase in the signal-to-noise requirements of surrogate matrix LLOQs were therefore not required, and allowed establishment of a 15 pg/mL LLOQ with signal-to-noise ≥ 5 .

The assay was not sensitive enough to detect endogenous concentrations of GLP-1, therefore it was quantified using an exogenous quantitation strategy, where both calibrants and QCs were prepared in the authentic plasma matrix.

Qualification

The FDA [24] and EMA [25] bioanalytical validation guidelines do not formally consider large molecule or biomarker analysis by LC-MS/MS. However they are often used as the basis for such studies, as in this current work, as many experiments are applicable and alternative regulatory guidelines do not exist. When setting acceptance criteria for biomarker assays it is important to consider the natural variability of the biomarker, for glucagon plasma concentrations can vary widely (3 - 130 pg/mL), and therefore a very stringent acceptance criteria was not considered necessary. Considering this and the analytical challenges associated with the surrogate nature of the assay, a 20% value was used for glucagon precision and accuracy across all levels, which is often used for immunoassays. For GLP-1 a 25% criteria at all levels was used for the interpretation of data from the dosed subjects because of the analytical challenges presented by the lack of a SIL or closely related analogue IS.

Glucagon

The performance of the developed method was initially assessed using formic acid based mobile phases. The assay was linear over the 15 - 1000 pg/mL calibration range assessed. Signal to noise was ≥ 5 at the LLOQ level (Figure 3a). Precision and accuracy (%RE and %CV) were within the 20% acceptance criteria across all levels assessed (Supplementary Information Table 5).

Matrix enhancement was observed, however this was compensated by the enhancement of the SIL IS, so it did not affect quantitation (Supplementary Information Table 6). The assay

gave acceptable performance across 6 individuals matrices fortified with 200 pg/mL of glucagon, and also in lithium heparin plasma providing further evidence that the assay was not significantly affected a matrix changes (Supplementary Information Table 5). There were no interferences in the selectivity samples near glucagon's retention time that could affect the quantitation. Recovery, determined at 200 pg/mL, was 50.7%.

The performance of the method was similarly assessed using mobile phases modified with *m*-NBA, and comparable performance was observed (Supplementary Information Table 5 and 6) demonstrating the methods suitability for providing a qualifier SRM to further assure selectivity as discussed later.

Glucagon was unstable at room temperature (65.9%) but stable on ice (110.1%) when incubated for 4 hr 50 minutes, consistent with previous observations [31], which sufficiently encompassed processing times for all batches extracted. The ability to re-inject extracts was demonstrated after 42 days storage at 4°C (Supplementary Information Table 11). Samples could be diluted 10-fold without affecting quantitation, enabling analysis in cases of low sample volume (Supplementary Information Table 5).

Figure 3 Glucagon and GLP-1 LLOQ chromatograms. a)Glucagon surrogate matrix sample at 15pg/mL (697.5/694.0) (S/N=7.6). b) GLP-1 plasma sample at 25 pg/mL (660.6/752.0) (S/N= 5.6).

GLP-1

A SIL IS or a closely related analogue was not available for GLP-1. The glucagon internal standards were therefore selected to investigate whether they offered advantages over IS free quantitation for GLP-1. The GLP-1 assay was characterised for precision and accuracy using two different SRM transitions for each mobile phase to ensure the most appropriate was selected.

Precision and accuracy (%RE and %CV) using formic acid phases was acceptable ($\leq 25\%$) for all but one analyte SRM and internal standard combinations, with the 660.6/752.0 GLP-1 SRM with the analogue glucagon internal standard giving the best overall sensitivity and performance (Supplementary Information Table 7). Signal to noise was ≥ 5 at the LLOQ level (Figure 3b). Performance was generally worse using *m*-NBA modified phases, especially at the LLOQ level. This is likely due to the glucagon internal standards performing differently under supercharging conditions to the GLP-1 analyte (Supplementary Information Table 7). Such phases were therefore not investigated further for GLP-1.

GLP-1 matrix suppression was observed, whereas the analogue glucagon IS experienced matrix enhancement, however the ratio between the two was similar across the matrices (CV= 16%), so there was no effect on quantitation (Supplementary Information Table 8). In addition the assay gave acceptable performance across 6 individual matrices, and 6 replicates of lithium plasma spiked with 200 pg/mL of analyte (Supplementary Information Table 7).

There were no interferences in the selectivity samples at the GLP-1 retention time. As expected [31] GLP-1 was unstable at room temperature (9.6%) but stable on ice for 4 hr 50 minutes (94.7%). The ability to re-inject extracts was demonstrated after 42 days storage at 4°C (Supplementary Information Table 11). GLP-1 recovery determined at 200 pg/mL (53.7%) was similar to glucagon's, and it was demonstrated that samples could be diluted 10-fold without affecting quantitation (Supplementary information Table 7).

Analysis of plasma from volunteers to determine endogenous glucagon concentrations

Plasma was collected from 11 healthy males and 11 healthy females and glucagon levels determined using the qualified LC-MS/MS method with standard formic acid mobile phases. Plasma glucagon concentrations were between <15 pg/mL (BLQ) and 47.1 pg/mL (Table 1), similar to ranges reported in the literature (~3 to 130 pg/mL [9] or ~27- 87 pg/mL [20]).

Several samples contained evidence of glucagon, but the concentrations were BLQ (Figure 4). Numerical values were determined for indicative purposes by extrapolation of the calibration line, and are shown in parenthesis in Table 1. The performance of the method at the 10 pg/mL level, as determined by evaluation of surrogate matrix QCs, was reasonable (mean accuracy 87.4%, CV= 29.7%), however poor performance resulted at the 7.5 pg/mL level (Mean accuracy 68.6%, CV= 40.4%). This should be considered whilst evaluating concentrations below the established range.

Figure 4 –Glucagon chromatograms from a) BLQ (14.7 pg/mL) and b) within range (40.0 ng/mL) sample

Samples with glucagon concentrations levels above the LLOQ were reanalysed to assess incurred sample reproducibility (ISR) (Table 1). 6/9 (66.6%) of samples gave concentrations within 20% of the original concentration, therefore passing standard criteria. A further sample gave a BLQ value upon reanalysis, although as its original concentration was close to the 15 pg/mL LLOQ, this was within the expected range. The storage time before reanalysis varied from 6 - 249 days, but no trend between the % difference and storage time was observed.

Similarly, a pooled plasma sample (n=6 replicates) gave a re-assayed concentration (33.3 pg/mL, CV=9.4%) within 4.3% of its original concentration (31.9 pg/mL, CV=6.1%) following additional storage for 42 days at -80°C (Supplementary information Table 9). This further demonstrated the method's good reproducibility at the endogenous glucagon level.

Table 1 Incurred sample reproducibility (ISR) assessment; comparison of original and reassayed endogenous glucagon concentrations following storage. NS- No sample remaining for repeat analysis. % difference= 100* (reassayed concentration- original concentration)/original concentration. Pass= % difference within 20%, Fail= %difference outside 20%.

Sample	Days stored (-80°C) before original analysis	Original Concentration (pg/mL)	Days stored at -80°C between analyses	ISR concentration (pg/mL)	% difference	Pass/Fail
M1	95	19.1	152	25.3	27.9	Fail
M2	74	BLQ (8.50)	152	-	-	-
M3	74	25.7	-	NS	-	-
M4	11	42.8	152	39.1	-9.0	Pass
M5	11	34.1	249	44.8	27.2	Fail
M6	11	27.2	249	27.8	2.2	Pass
M7	15	BLQ (12.5)	-	-	-	-
M8	36	BLQ (No Peak)	-	-	-	-
M9	36	BLQ (No Peak)	-	-	-	-
M10	38	BLQ (14.1)	-	-	-	-
M11	38	45.9	6	50.4	9.3	Pass
F1	95	BLQ (14.7)	-	-	-	-
F2	95	40.0	152	41.9	-4.6	Pass
F3	74	15.1	152	15.7	3.9	Pass
F4	11	33.8	152	23.4	-36.4	Fail
F5	11	BLQ (No Peak)	-	-	N/A	-
F6	10	BLQ (10.3)	-	-	-	-
F7	15	BLQ (9.46)	-	-	N/A	-
F8	15	26.9	99	24.7	-8.5	Pass
F9	36	BLQ (14.7)	-	-	-	-
F10	38	BLQ (14.5)	-	-	-	-
F11	38	15.5	6	BLQ (No Peak)	-	NA

For exogenous compounds the absence of significant peaks at the analyte retention time can be used to demonstrate the selectivity of the method. This is not possible for endogenous compounds, and therefore additional SRM transitions can be monitored as qualifier transitions to ensure selectivity [16][17][18][19]. However for glucagon the only SRM sensitive enough to determine endogenous concentrations using formic acid mobile phases was that monitored (697.5/694.0). The method with mobile phases modified with *m*-NBA enabled access to an alternative sensitive qualifier SRM transition (581.5/578.5), derived from a higher charge state, as well as providing alternative LC conditions. Reanalysis of extracts using this method demonstrated concentrations were not significantly different to those originally obtained providing additional evidence for the selectivity of the formic acid

based mobile phase method (Table 2). As far as the authors knowledge the use of supercharging mobile phases for such an application has not been reported.

Table 2 Determination of endogenous plasma gluagon concentrations using formic acid and *m*-NBA modified mobile phases and corresponding SRM transitions % difference= 100*

(Qualification assay concentration – quantitative assay concentration)/quantitative assay concentration. BLQ. Below the limit of quantitation (BLQ). Extrapolated concentrations given in brackets. NC- not calculatable.

Sample	Glucagon Concentration (pg/mL)		% Difference
	Quantitative assay	Qualification assay	
	0.2% FA Phases Glucagon (697.5/694.0)	0.05% <i>m</i> -NBA 0.01% FA phases Glucagon (581.5/578.5)	
F1	BLQ (14.7)	BLQ (12.6)	-14.3
F2	40.0	43.3	8.2
F3	15.1	17.2	13.9
F4	33.8	32.8	-3.0
F5	No Peak	No Peak	NC
M1	19.1	23.1	20.9
M2	BLQ (8.50)	No Peak	NC
M3	25.7	26.8	4.3
M4	42.8	46.2	7.9

Cross validation against established immunoassays using physiological samples

Plasma samples were available to cross-validate the LC-MS/MS method against established immunoassays. These originated from a physiological study assessing the preanalytical stability of glucagon and GLP-1 in the presence of various stabilisers and anticoagulants. These did not introduce a significant matrix effect affecting LC-MS/MS quantitation (Supplementary Information Table 10). Samples originated from fasting healthy volunteers (endogenous glucagon), volunteers dosed with glucagon, and volunteers dosed with GLP-1. Results of the stability investigations at the endogenous glucagon level have been recently reported [32]. Two immunoassays using different assay formats were used to cross-validate each analyte against the LC-MS/MS method, to help provide confidence, or otherwise, of the immunoassays results.

For endogenous glucagon quantitation, the HTRF immunoassay demonstrated limited utility as only 16/50 samples were above the LLOQ (12.0 pg/mL). In contrast, the majority of the samples analysed by the Milliplex assay (38/45) and LC-MS/MS method (39/50) gave quantifiable ($\geq$ LLOQ) concentrations. Whilst this may be expected for the LC-MS/MS method, with its lower LLOQ (10.0 pg/mL), the Milliplex had a higher LLOQ (13.0 pg/mL). This discrepancy may be due to differences in the standardisation of reference materials between the immunoassays and/or the Milliplex being more susceptible to interferences. Only data above the LLOQ is discussed below, as large errors would likely be associated with lower concentrations, although for completeness all concentrations determined are shown in Supplementary Information Figure 11a.

The endogenous glucagon concentrations measured by all three methods were within literature ranges (~3 to 130 pg/mL [9] or ~27- 87 pg/mL [20]) (Figure 5 and Supplementary Figure 11b). Where both the HTRF and LC-MS/MS methods gave quantifiable ($\geq$ LLOQ) concentrations, these were of similar magnitude with no notable inter-volunteer or anticoagulant effects (Figure 5a). Likewise the LC-MS/MS and Milliplex methods gave similar concentrations for samples from volunteers 1, 2, and 5, and from the EDTA samples from volunteer 3 (Figure 5b). However the Milliplex gave significantly higher concentrations for lithium heparin samples obtained from volunteer 3, and for all samples from volunteer 4. The Milliplex assay also showed a positive bias for samples from volunteer 4 in comparison to the HTRF when quantifiable ($\geq$ LLOQ) results were compared (Figure 5c). This suggests that the matrix from volunteer 4 affects Milliplex quantitation due to the presence of low level interferences sharing an epitope with glucagon or due to other non-specific interferences, as has been suggested previously [9]. Similarly, interferences may be present in lithium heparin samples from volunteer 3, affecting immunoassay quantitation. The poor sensitivity of the

HTRF, and potential susceptibility of the Milliplex assay to matrix interferences, demonstrates some of the challenges in selecting appropriate immunoassays for LC-MS/MS cross validation.

Figure 5 Comparison of endogenous glucagon concentrations determined by LC-MS/MS and two immunoassays. a) HTRF vs LC-MS/MS. b) Milliplex vs. LC-MS/MS, c) HTRF vs. Milliplex. Samples were analysed from 5 volunteers (1-5), which contained either EDTA (E) or Lithium Heparin (L) anticoagulant in addition to various stabilisers (not shown). Different stabiliser combinations were used in samples of the same Sample Type, e.g. the four 4E samples in Figure 5b are all unique. Only samples with quantifiable ($\geq$ LLOQ) results for both assays being compared are displayed. Full results are shown in Supplementary Information Figure 11.

Concentrations of the samples from the volunteer dosed with glucagon strongly correlated between the two immunoassays, and similar absolute concentrations were determined (Figure 6a). This may be due to the absence of matrix interferences affecting the Milliplex assay from this volunteer, or interferences becoming negligible at higher glucagon concentrations. The LC-MS/MS method showed some similarity in concentrations trends between samples, although results were less comparable, and absolute concentrations were generally higher. This may reflect differences in the natures of the techniques; immunoassays detect specific epitopes whereas the LC-MS/MS method targets the whole molecule. It has been recently suggested that epitope masking caused by glucagon aggregation may artificially lower immunoassay assay concentrations in plasma samples fortified with high levels of glucagon [26]. A similar process may occur in samples from dosed patients also containing high levels of glucagon. In addition to generally higher LC-MS/MS concentrations, this could also explain differences in concentration trends if aggregation occurred to various degrees in different samples.

Figure 6 Comparison of concentrations determined by LC-MS/MS and two immunoassays in samples from volunteers dosed with a) glucagon and b) GLP-1. Samples were analysed from 1 volunteer (labelled 6) dosed with glucagon and 2 volunteers (labelled 7 & 8) dosed with GLP-1. Samples contained either EDTA (E) or Lithium Heparin (L) anticoagulant in addition to various stabilisers (not shown). Different stabiliser combinations were used in samples of the same Sample Type, e.g the six 6E samples are all unique.

Samples from the two volunteers dosed with GLP-1 showed similar concentrations trends between the samples across the three methods (Figure 6b). Absolute concentrations showed a systematic bias of LC-MS/MS > Millipore > Milliplex. This could be due to differences in the standardisation of reference materials. Although it was also noted that lithium heparin samples from subject 7 showed similar immunoassay concentrations, suggesting a susceptibility of the assays to inter-volunteer and anticoagulant related matrix effects. The limited agreement of the immunoassays restricted their value to further cross-validate the LC-MS/MS method.

Overall, the cross-validation provided only limited confirmation for the performance of the LC-MS/MS method. However this was not unexpected as the poor performance of some immunoassays provided the main rationale for development of the LC-MS/MS method. Further investigations are needed with larger data sets, various sample types, and additional assays, to better characterise the performance of the developed LC-MS/MS method.

635 **Conclusion**

636 We describe a multiplexed LC-MS/MS method for the quantitation of endogenous glucagon,
637 dosed glucagon, and/or dosed GLP-1 in human plasma without immunoenrichment. The
638 sensitivity achieved is the highest reported for both peptides for such an extraction. The
639 methodology has been improved over the initial method through a significant shortening of
640 extraction time, inclusion of a stable isotope labelled internal standard and a reduction in LC-
641 MS cycle time. The one day extraction compares well against many immunoassays, which
642 often incur overnight incubation steps, as required by the Millipore and Millipex assays used
643 in this study. Samples extracted for LC-MS/MS analysis can be run overnight to produce
644 data in a similar overall time frame to immunoassays. Good precision and accuracy was
645 obtained over the calibration range, and recovery, matrix effects, and bench top stability
646 were acceptable. Endogenous glucagon specificity was assured by a novel approach using
647 mobile phases modified with the *m*-NBA supercharging additive to access an alternative
648 sensitive qualifier SRM transition.

649 Endogenous glucagon concentrations were within the expected range, and determination
650 was highly reproducible after extended sample storage (up to 249 days at -80°C). A cross-
651 validation was performed using physiological samples with the qualified LC-MS/MS method
652 and two immunoassays for each analyte. Some similarities between the LC-MS/MS method
653 and immunoassays were observed for endogenous glucagon, but dosed glucagon and
654 dosed GLP-1 samples showed a trend towards higher concentrations when measured by
655 LC-MS/MS. It was however noted that immunoassays also showed some differences
656 between each other, limiting their value for cross validation. Such differences were not
657 wholly unexpected, as performance concerns associated with glucagon and GLP-1
658 immunoassays provided the main rationale for development of the LC-MS/MS method. We
659 therefore consider the LC-MS/MS method to be a viable alternative to immunoassay based
660 approaches.

661 However we appreciate that cost may be a limitation to the adoption of LC-MS/MS to routine
662 clinical diagnostics in some laboratories. Whilst consumable costs are lower, with the largest
663 being the SPE plate at a couple of hundred pounds (compared to several hundred pounds
664 for immunoassays kits), LC-MS/MS instrumentation is considerably more expensive.
665 Sensitive LC-MS/MS systems cost in the region of two hundred thousand pounds, whereas
666 plate readers can be obtained for around 10 times less. Although as a platform technology
667 LC-MS/MS enables rapid amortisation of the capital cost over multiple analytes. The uptake
668 of LC-MS/MS within clinical laboratories is increasing, and therefore there is the potential to

initiate specialist centres with access to LC-MS/MS instrumentation, who could set up and run a glucagon LC-MS/MS service.

Future Perspectives

As instrumentation improves it will be possible to accurately quantify lower levels of endogenous glucagon as well as endogenous GLP-1 without the use of immunoenrichment. These will also allow sensitive qualifier SRM transitions to be obtained using standard formic acid mobile phases to ensure endogenous compound specificity, until then transitions may be accessed using mobile phase supercharging additives as described. The use of high throughput 2D extractions to minimise matrix suppression, may also become a more widespread strategy to improve sensitivity. Further development could also include the expansion of the assay to other proglucagon derived peptides, or other peptides of interest. Reproduction of LC-MS/MS methods at different laboratories will provide further confidence of the performance and robustness of these.

The differences in glucagon and GLP-1 concentrations obtained using the various analytical methods is concerning, as they could lead to incorrect clinical decisions. Assays should be better characterised with regards to inter-subject and anticoagulant related matrix interferences, as well as for cross-reactivity with metabolites and degradation fragments. Characterisation may include the use of biological matrix reference materials containing independently certified amounts of analyte [33], but these are not currently available for glucagon or GLP-1. The ability of glucagon to aggregate in plasma, and its effect on quantitation, requires further study, and the standardisation of reference materials between assays should be improved. The benefit of using LC-MS/MS based methodologies is that specificity is enhanced over immunoassays, as they can target the full length molecule rather than specific epitopes, and therefore such methods may become increasingly utilised in the future.

Executive Summary

Background

- Glucagon and GLP-1 are traditionally quantified in human plasma by immunoassay. However precision and accuracy can be poor, as can be the correlation between assays.
- LC-MS/MS based approaches can be used as alternatives to help overcome such challenges; however their application has been limited by lower sensitivity.
- This study refined and expanded our previous glucagon-only method, improving sensitivity, robustness, and throughput, and further demonstrated specificity of endogenous glucagon quantitation. GLP-1 was also introduced as a secondary analyte.

Experimental

- Method refinements included modification of the 2D (protein precipitation then SPE) extraction protocol and the introduction of an automated liquid handling system to improve sensitivity, robustness, and throughout.
- The multiplexed method was qualified based on key experiments described in the latest FDA and EMA bioanalytical guidelines.
- A sensitive qualifier SRM was accessed via a novel application of a supercharging mobile phase additive to ensure specificity of endogenous glucagon quantitation.
- Physiological study samples were quantified using the qualified LC-MS/MS assay, and a cross-validation was performed against established immunoassays.

Results and Discussion

- Method refinements halved the extraction time to 1 working day (96 samples) and UHPLC run time to 3.6 minutes/samples, significantly improving throughput. Sensitivity was improved by reduced matrix suppression in the refined method.
- The method was successfully multiplexed with GLP-1, and met the qualification criteria for both analytes.
- Endogenous glucagon concentrations were determined within the expected range, and confirmed by comparison to a sensitive qualifier SRM. Concentrations showed good reproducibility after extended sample storage.
- Glucagon and GLP-1 concentrations were determined from physiological study samples containing endogenous glucagon, dosed glucagon and dosed GLP-1.
- Cross-validation against immunoassays using the physiological study samples showed some similarities between the methods. However differences were not wholly unexpected, as performance concerns associated with glucagon and GLP-1 immunoassays provided the main rationale for development of the LC-MS/MS method.

Conclusion

- The sensitivity achieved (Glucagon- 15 pg/mL LLOQ, GLP-1- 25 pg/mL LLOQ) is the highest reported for both peptides for an extraction avoiding immunoenrichment
- The method provides a viable alternative to immunoassay based approaches for the quantitation of endogenous glucagon, dosed glucagon, and/or dosed GLP-1 from human plasma samples.

Financial & competing interests disclosure

The authors have no relevant affiliations or financial involvement with any organization or entity with a financial interest in or financial conflict with the subject matter or materials discussed in the manuscript. This includes employment, consultancies, honoraria, stock ownership or options, expert testimony, grants or patents received or pending, or royalties.

No writing assistance was utilized in the production of this manuscript.

Ethical conduct of research

The authors state that they have obtained appropriate institutional review board approval or have followed the principles outlined in the Declaration of Helsinki for all human or animal experimental investigations. In addition, for investigations involving human subjects, informed consent has been obtained from the participants involved.

References of interest

[10]—Comprehensive overview of immunological approaches for the quantitation of the various GLP-1 isoforms, including practical considerations.

[28] Description of the mechanism via which supercharging molecules lead to increased analyte charging.

[30]- Evaluation of various strategies for endogenous quantitation.

References of considerable interest

[9] - All seven glucagon immunoassays evaluated demonstrated poor performance in terms of specificity, spiked recovery, susceptibility to matrix effects/non-specific matrix interferences and/or and sensitivity.

[11]- Considerable variability in specificity and sensitivity was observed between the ten GLP-1 immunoassays kits evaluated.

[20] –An Immunoaffinity extraction with microflow LC was used to develop the most sensitive glucagon (2.7 pg/mL LLOQ) and GLP-1 (2.6 pg/mL LLOQ) LC-MS/MS method in the literature.

References

- [1] Sandoval D a, D'Alessio D a. Physiology of Proglucagon Peptides: Role of Glucagon and GLP-1 in Health and Disease. *Physiol. Rev.* 95(2), 513–548 (2015).
- [2] Kolb A, Rieder S, Born D, et al. Glucagon/insulin ratio as a potential biomarker for pancreatic cancer in patients with new-onset diabetes mellitus. *Cancer Biol. Ther.* 8(16), 1527–1533 (2009).
- [3] Furuse M, Matsumoto M, Okumura JI, Sugahara K, Hasegawa S. Intracerebroventricular injection of mammalian and chicken glucagon-like peptide-1 inhibits food intake of the neonatal chick. *Brain Res.* 755(1), 167–169 (1997).
- [4] Holst JJ. The Physiology of Glucagon-like Peptide 1. *Physiol. Rev.* (225), 1409–1439 (2007).
- [5] De Graaf C, Blom WAM, Smeets PAM, Stafleu A, Hendriks HFJ. Biomarkers of satiation and satiety. *Am. J. Clin. Nutr.* 79(6), 946–961 (2004).
- [6] Parker J a, McCullough K a, Field BCT, et al. Glucagon and GLP-1 inhibit food intake and increase c-fos expression in similar appetite regulating centres in the brainstem and amygdala. *Int. J. Obes. (Lond).* 37(10), 1391–8 (2013).
- [7] Tan TM, Field BCT, McCullough K a, et al. Coadministration of glucagon-like peptide-1 during glucagon infusion in humans results in increased energy expenditure and amelioration of hyperglycemia. *Diabetes.* 62(4), 1131–8 (2013).
- [8] Alford FP, Bloom SR, Nabarro JD. Glucagon levels in normal and diabetic subjects: Use of a specific immunoabsorbent for glucagon radioimmunoassay. *Diabetologia.* 13(1), 1–6 (1977).
- [9] Bak MJ, Albrechtsen NW, Pedersen J, et al. Specificity and sensitivity of commercially available assays for glucagon and oxyntomodulin measurement in humans. *Eur J Endocrinol.* 170(4), 529–38 (2014).
- [10] Kuhre RE, Wewer Albrechtsen NJ, Hartmann B, Deacon CF, Holst JJ. Measurement of the incretin hormones: Glucagon-like peptide-1 and glucose-dependent insulinotropic peptide. *J. Diabetes Complications.* 29(3), 445–450 (2015).
- [11] Bak MJ, Albrechtsen NJW, Pedersen J, et al. Specificity and sensitivity of commercially available assays for glucagon-like peptide-1 (GLP-1): implications for GLP-1 measurements in clinical studies. *Diabetes, Obes. Metab.* 16, 1155–1164 (2012).
- [12] Heijboer AC, Frans A, Lomecky M, Blankenstein MA. Analysis of glucagon-like peptide 1; what to measure? *Clin. Chim. Acta.* 412(13–14), 1191–4 (2011).
- [13] Veniamin N Lapko, Patrick S Miller, G Paul Brown, Rafiqul Islam, Sarah K Peters, Richard L Sukovaty PFR& CJK. Sensitive glucagon quantification by immunochemical and LC – MS / MS methods. *Bioanalysis.* 5(23), 2957–2972 (2013).
- [14] Ørskov C, Rabenhøj L, Wettergren A, Kofod H, Holst JJ. Tissue and Plasma Concentrations of Amidated and Glycine-Extended Glucagon-Like Peptide I in Humans. *Diabetes.* 43(4), 535–539 (1994).
- [15] Howard JW, Kay RG, Tan T, Minnion J, Creaser CS. Identification of plasma protease derived metabolites of glucagon and their formation under typical laboratory

812 sample handling conditions. *Rapid Commun. Mass Spectrom.* 29(2), 171–181 (2015).

813 [16] Honour JW. Development and validation of a quantitative assay based on tandem
814 mass spectrometry. *Ann Clin Biochem.* 48(2), 97–111 (2011).

815 [17] Abbatiello SE, Mani DR, Keshishian H, Carr SA. Automated detection of inaccurate
816 and imprecise transitions in peptide quantification by multiple reaction monitoring
817 mass spectrometry. *Clin. Chem.* 56(2), 291–305 (2010).

818 [18] Holman SW, Sims PFG, Evers CE. The use of selected reaction monitoring in
819 quantitative proteomics. *Bioanalysis.* 4(14), 1763–86 (2012).

820 [19] Addona TA, Abbatiello SE, Schilling B, et al. Multi-site assessment of the precision
821 and reproducibility of multiple reaction monitoring-based measurements of proteins in
822 plasma. *Nat. Biotechnol.* 27(7), 633–41 (2009).

823 [20] Lee A, Chappell D, Bak M, et al. Multiplexed Quantification of Proglucagon-Derived
824 Peptides by Immunoaffinity Enrichment and Tandem Mass Spectrometry after a Meal
825 Tolerance Test. *Clin Chem.* 62(1), 227–35 (2016).

826 [21] Howard JW, Kay RG, Tan T, et al. Development of a high throughput UHPLC-MS/MS
827 (SRM) method for the quantitation of endogenous glucagon from human plasma.
828 *Bioanalysis.* 6(24), 3295–3309 (2014).

829 [22] Chappell D, Lee A, Castro-Perez J, et al. An ultrasensitive method for the
830 quantitation of active and inactive GLP-1 in human plasma via immunoaffinity LC-
831 MS/MS. *Bioanalysis.* 6(1), 33–42 (2014).

832 [23] Amao M, Kitahara Y, Tokunaga A, Shimbo K, Eto Y, Yamada N. Simultaneous
833 quantification of intracellular and secreted active and inactive glucagon-like peptide-1
834 from cultured cells. *Anal. Biochem.* 472, 45–51 (2015).

835 [24] *Guidance for industry: Bioanalytical method validation.* U.S. Department of Health and
836 Human Services, Food and Drug Administration, Center for Drug Evaluation and
837 Research (CDER), Center for Veterinary Medicine (CVM), May 2001, (2001).

838 [25] Guideline on bioanalytical method validation, EMA(2012).

839 [26] Wewer Albrechtsen NJ, Bak MJ, Hartmann B, et al. Stability of glucagon-like peptide
840 1 and glucagon in human plasma. *Endocr. Connect.* 4(1), 50–57 (2015).

841 [27] Kay RG, Howard J, Stensson S. A current perspective of supercharging reagents and
842 peptide bioanalysis. *Bioanalysis.* 8(3), 157–161 (2016).

843 [28] Iavarone AT, Williams ER. Mechanism of charging and supercharging molecules in
844 electrospray ionization. *J. Am. Chem. Soc.* 125(8), 2319–27 (2003).

845 [29] Houghton R, Horro Pita C, Ward I, Macarthur R. Generic approach to validation of
846 small-molecule LC-MS/MS biomarker assays. *Bioanalysis.* 1(8), 1365–74 (2009).

847 [30] Jones BR, Schultz G a, Eckstein J a, Ackermann BL. Surrogate matrix and surrogate
848 analyte approaches for definitive quantitation of endogenous biomolecules.
849 *Bioanalysis.* 4(19), 2343–56 (2012).

850 [31] Deacon C, Johnsen A, Holst J. Degradation of glucagon-like peptide-1 by human
851 plasma in vitro yields an N-terminally truncated peptide that is a major endogenous
852 metabolite in vivo. *J Clin Endocrinol Metab.* 80(3), 925–7 (1995).

- 853 [32] Cegla J, Jones BJ, Howard J, et al. The preanalytical stability of glucagon as
854 measured by liquid chromatography tandem mass spectrometry and two
855 commercially available immunoassays. *Ann. Clin. Biochem.*(2016).
- 856 [33] French D. Development and validation of a serum total testosterone liquid
857 chromatography–tandem mass spectrometry (LC–MS/MS) assay calibrated to NIST
858 SRM 971. *Clin. Chim. Acta.* 415, 109–117 (2013).
- 859

1 ***Bioanalysis* 2017, Vol. 9 (9), pp. 733-751**

2 Development of a UHPLC-MS/MS (SRM) method for the quantitation of endogenous
3 glucagon and dosed GLP-1 from human plasma

4 **Supplementary Information**

5 **Table S1** Parameters re-optimised during assay development for glucagon quantitation

Parameter Re-optimised		Aim	Result	Method altered?	Sup Info Ref
Protein Precipitation Solvent	Composition	Increase sensitivity by increasing recovery and/or reducing matrix effects	Original solvent (75/25/0.1, ACN/H ₂ O/NH ₃) gave optimal performance	No	1.1
	Volume	Reduce extraction time by reducing solvent volume to enable use of 96 well plates and associated automated liquid handling, as well as reducing evaporation time.	Volume reduced from 3.2 mL to 1.1 mL. Automated liquid handling utilised. Extraction time halved to 1 day.	Yes	1.2
	Evaporation temperature	Reduce evaporation time	Higher temperatures (60°C) led to increased matrix interference.	No	1.3
Extract Reconstitution Volume		Increase sensitivity by reducing reconstitution volume to increase extract concentration	Reduced volume increases background noise, lowering signal-to-noise.	No	2
LC	Flow Rate	Increase sensitivity, as a result of higher ionisation efficiencies at lower flow rates. Monitor multiple charge states to ensure optimal is selected.	No overall boost in sensitivity was observed due to increased peak broadening at lower flow rates. Original charge state was optimal.	No	3.1
	Formic acid composition	Increase sensitivity, by increasing positive electrospray ionisation at lower pH	Sensitivity <i>reduces</i> at lower pH.	No	3.2
	Column clean time	Reduce run time by reducing column clean, enabled by lower matrix build up by automated liquid handling.	Column clean time reduced by 3 minutes, leading to a gradient time of 3.6 min per sample, saving 5 hours of analysis time per 96 sample batch.	Yes	3.3
Internal Standard		Increase precision and accuracy, and robustness using SIL rather than an analogue glucagon IS	SIL IS gave better performance under conditions of high matrix effect.	Yes	2 & 4

1 Extraction Optimisation

1.1 Protein Precipitation Solvent Composition

Previously we investigated ACN based protein precipitation solvents diluted to various proportions (0, 25, 50%) with water [1]. We found that 75/25 ACN/H₂O gave the best response, and that the addition of 0.1% formic acid significantly reduced it, whilst the addition 0.1% ammonia gave a small increase. Therefore 75/25/0.1 ACN/H₂O/NH₃ was selected. To attempt to increase performance further in this study we investigated additional proportions of ACN in water (20, 15, 10, 5%) and investigated the use of MeOH based solvents, however the original solvent remained optimal (Figure S1). We have found in our laboratory that using 25% water in acetonitrile is often the optimal protein precipitation solvent for peptide analysis.

Figure S1 Effect on analyte and analogue internal standard (IS) peak areas upon extracting plasma samples (100 μ L) spiked with 1 ng/mL glucagon with various protein precipitation solvents

1.2 Protein Precipitation Solvent Volume

The protein precipitation solvent volume was reduced (3.2 mL to 1.1 mL) to enable the assay to be performed in a 2mL 96 well plate, rather than tubes. Interference peaks (1.6 min) close to glucagon's retention time (1.7 min) were eliminated (Figure S2 a vs b).

Figure S2 Effect of the extraction method on interference peaks. The matrix interference (1.6 min) close to the analyte retention time observed using the a) original extraction procedure is removed using the b) modified extraction procedure. c) Increasing the solvent evaporation temperature to 60°C resulted in a slight decrease in internal standard response and an increase in tailing.

1.3 Protein Precipitation Solvent Evaporation Temperature

The blow down temperature was increased from 40°C to 60°C to attempt to reduce evaporation time. No significant difference in analyte signal (Figure S2 b vs c) or QC accuracies were observed, however the IS showed slightly lower signal and greater tailing. As the difference in evaporation time was small (75 min vs 95 min), it was decided to maintain the temperature at 40°C.

2 Extract Reconstitution Volume

Lower extract reconstitution volumes were investigated to attempt to boost sensitivity. However whilst signal did increase the background noise increased to a greater extent, leading to a reduction in signal-to-noise with a constant 40 µL injection volume (18.4, 15.5, 12.9, at 200, 150, 100 µL reconstitution volumes respectively for a 100 pg/mL sample).

The Analyte/internal standard peak-area-ratio (PAR) was also affected at low reconstitution volumes (100 µL) using the analogue internal standard for 100 and 750 pg/mL samples, affecting quantitative accuracy and demonstrating the limitation of the analogue internal standard under conditions of high matrix effects (Table 2).

Table 2 Effect of reconstitution volume on extracted plasma QC performance, quantified using the analogue and SIL glucagon internal standard.

Reconstitution volume (µL)	Glucagon Internal Standard							
	Analogue				SIL			
	100 pg/mL		750 pg/mL		100 pg/mL		750 pg/mL	
	% 200 µL PAR	%CV	% 200 µL PAR	%CV	% 200 µL PAR	%CV	% 200 µL PAR	%CV
200	100.0	2.0	100.0	5.7	100.0	1.6	100.0	4.1
150	97.7	1.7	99.8	0.2	103.7	1.6	102.8	1.1
100	78.1	3.6	89.8	6.9	98.5	1.3	104.3	2.6

3 LC Optimisation

3.1 Flow Rate

Lower LC-flow rates (0.25 mL/min and 0.40 mL/min) were investigated to see whether they increased sensitivity over the original (0.80 mL/min) flow rate, due to improved ionisation efficiency. To reduce the effect of peak broadening flow rates were lowered only during analyte elution.

Lower flow rates were noted to increase the relative formation of the highly charged 4+ ion ($m/z = 697$), as expected due to prolonged interaction with the electrospray probe allowing greater charge accumulation. Peak areas increased with lower flow rates (Figure S3b), however peak broadening was still significant, and therefore there was no overall significant increase in peak height (sensitivity) (Figure S3b). The flow rate was maintained at 0.8 mL/min.

Figure S3 Effect of the flow rate on (a) peak area and (b) heights for various glucagon charge states using 1ng/mL solution. Flow rate is 0.8 mL/min except at elution. Mobile phases A= 0.2% FA (ACN) B= 0.2% FA (aq).

3.2 Formic Acid Composition

The percentage of formic acid in the mobile phases was investigated using 1 ng/mL glucagon solutions (Figure S4). Reducing formic acid from 1% to 0.2% increased the intensity of all charge states, including the most highly charged 5+ ($m/z = 697$) charge state, without significantly affecting the charge state distribution. Higher percentages of formic acid may increase ionisation of matrix components, limiting ion current available for analyte ions. However the proportion of the 5+ ($m/z = 697$) charge state reduced in abundance when formic acid was lowered further to 0.1%, as the lower 4+ ($m/z = 871$) charge state was favoured. Peaks height and peak area were similarly affected, as there was not effect on peak broadening. 0.2% FA remains optimal for the 5+ ($m/z = 697$) charge state.

Figure S4 Effect of the formic acid content on glucagon (a) peak area and (b) height. Flow rate is 0.8 mL/mL, mobile phases are A=ACN and B=water, each modified with formic acid to the extent shown in the figure.

3.3 Column clean time

The original LC method included a 4 minute column clean step at the end of each injection at the LC starting conditions (22% organic). This was found necessary to reduce column fouling, and resulted in a run time of 7.1 minutes. To attempt to reduce run time this was reduced to 0.4 minutes, which as expected led to column fouling during the injection of samples extracted using the original extraction method (Figure S5a vs b), however peak

shape was maintained using the re-optimised extraction method with automated liquid transfer (Figure S5c vs d). This was attributed to the reduction of the inadvertent transfer of protein precipitate solid, and subsequent reduction of matrix build up on the column. The total run time for the final multiplexed method was 3.6 minutes, saving 5 hours of analysis time per 96 sample batch.

Figure S5 Comparison of SIL internal standard peak shape using the original extraction procedure at plasma injection a) 9 and b) 33, and using the re-optimised extraction procedure at plasma injection a) 9 and b) 33.

4 Internal Standard

Table 3 Performance of the analogue and SIL glucagon internal standard against plasma QC samples extracted using the original [1] or re-optimised extraction procedure with plasma samples

Plasma QC Concentration (pg/mL)	Extraction Method	Glucagon Internal Standard			
		Analogue		SIL	
		% Accuracy	%CV	% Accuracy	%CV
MED 100 pg/mL	Original	114.0	6.2	97.8	4.3
	Re-optimised	98.5	5.5	92.0	5.5
HIGH 750 pg/mL	Original	114.8	2.9	97.6	3.6
	Re-optimised	102.9	3.5	100.4	4.5

5 Mobile Phase Supercharging Reagents

5.1 Initial Optimisation

5.1.1 Optimisation of *m*-NBA content in mobile phases

The percentage of *m*-NBA in the mobile phases was found to have a large effect on the charge state distribution of glucagon (Figure S6). As expected higher percentages favoured the most highly charged species (6+, m/z =581). At low levels of *m*-NBA (0 - 0.50%) the response of both the (6+, m/z =581) and 6+ (m/z = 697) ions increased with *m*-NBA concentration, presumably due to an increase in overall ionisation efficiency. Whereas at higher levels of *m*-NBA the increase of the 6+ (m/z =581) charge state was at the expense of the others. Data points were only collected for the 3+ and 4+ ions at 0 and 0.1% *m*-NBA concentrations, as these ions were least intense and therefore discarded early on in method development.

Figure S6 The effect of *m*-NBA mobile phase content on charge state distribution. Mobile phases are A= ACN, B= H₂O/ACN (95/5), both modified with *m*-NBA as shown. A 1000 pg/mL surrogate standard (n=3 replicates) was analysed in each case. These data are from 3 different experiments (Exp1 (full scan): 0 & 0.1% *m*-NBA, Exp2 (SRM): 0.025 & 0.050% *m*-NBA, Exp 3 (SRM): 0.05 & 0.1% *m*-NBA). SRM transitions were 697/693 and 581/575. Data has been scaled appropriately, using common conditions between experiments, to account for signal variations.

5.1.2 Optimisation of Formic acid content in mobile phases

The effect of adjusting the percentage of formic acid in mobile phases was relatively small in standard mobile phases (Section 3.2), however when *m*-NBA was present the effect was significant (Figure S7). Acids are known to decrease the surface tension of electrospray droplets[2], and therefore altering concentrations could affect the supercharging process, which relies on increases in surface tensions by *m*-NBA [2].

Figure S7 The effect of formic acid content on glucagon SRM peaks area using *m*-NBA based mobile phases. Mobile phases are A= ACN, B= H₂O/ACN (95/5), both modified with either (a) 0.05% *m*-NBA with at 0, 0.010, 0.025, 0.050, or 0.0100% FA, or (b) 0.01% *m*-NBA with at 0, 0.010, 0.025, 0.0100, or 0.500% FA. Values are means of n≥3 replicate injections of 1000 pg/mL surrogate matrix standards.

5.2 Effect of *m*-NBA modified mobile phases upon analysis of GLP-1 and Glucagon Solutions and Plasma Extracts

A SRM method was created with 13 transitions (8 GLP-1 transitions and 5 for Glucagon or Glucagon IS) (Figure S8 and Figure S9). The SRM transitions encompassed the 5+ and 6+ parent ions for glucagon, and the 3, 4, 5 and 6+ parent ions for GLP-1. Product ions were identified by fragmentation of these parent ions, followed by collision energy optimisation. Multiple transitions were monitored to increase the likelihood of finding those with low background noise (thereby increasing signal to noise) along with the absence of interferences. A greater number of GLP-1 transitions were selected, as glucagon SRM transitions had been previously investigated.

A 20ms dwell time was used to ensure sufficient data points across the peaks. Scheduled MRM was used to maximise dwell time by only monitoring glucagon and GLP-1 transitions near their RT (1.6 and 2.5 respectively), a target scan time of 0.2 seconds, and MRM detection window of 30 seconds was selected.

Surrogate matrix and extracted plasma samples spiked to the 1200 pg/mL level with glucagon and GLP-1 were analysed using mobile phase modified with 0.05% or

0.01% *m*-NBA. These percentages were found to give the most intense analyte signal when glucagon solutions were analysed (Section 5.1.1). The mobile phases also contained either 0.01% FA, found to be optimal for glucagon solution analysis (Section 5.1.2) or 0.1% FA. The higher acid concentration was evaluated to investigate whether this was needed for effective analyte ionisation due to competition from matrix components.

The effect of altering mobile phases was similar for surrogate matrix (solution) samples and plasma extracts for both glucagon (Figure S8) and GLP-1 (Figure S9). In most cases it was noted that plasma extracts gave a lower signal, although this will at least in part be due to the loss of compound during the extraction, rather than difference in ionisation between the matrices. Mobile phases containing 0.05% and *m*-NBA 0.01% FA were found to give maximum signal, regardless of transition, nature of sample (plasma or solution), and analyte. The presence of matrix therefore did not alter the supercharging process. It was noted that some transitions were affected by the alteration in mobile phases more than others; demonstrating that supercharging of the analyte alters the intensity of various product ions formed.

Figure S8 Glucagon peak areas for 1200 pg/mL solution and plasma extracts under different mobile phase conditions. 1200 pg/mL solution or plasma extracts were analysed, values are means of n=3 replicates.

Figure S9 GLP-1 peak areas for 1200 pg/mL solution and plasma extracts under different mobile phase conditions. 1200 pg/mL solution or plasma extracts were analysed, values are means of n=3 replicates.

5.3 Comparisons of the optimised *m*-NBA modified mobile phase with the original formic acid mobile phase method

A plasma sample was spiked with 1200 pg/mL glucagon and GLP-1, and analysed with optimal *m*-NBA based mobile phase method (modified with 0.05% *m*-NBA 0.01% FA) and the original formic acid based method (modified with 0.2% FA). The signal-to-noise was assessed for each transition monitored (Table 4).

The transitions highlighted were selected for further assessments using either standard mobile phases (blue) or supercharging mobile phase (red). Although these were not always the transitions that gave the best S/N, they gave best overall performance when linearity and precision and accuracy were considered.

The SRM method was subsequently altered to remove redundant transitions, and split into separate methods for method for analysis with formic acid and supercharging mobile phase methods to maximise dwell time for each transitions to improve sensitivity. In addition scheduled MRM was replaced with periods, as this was found to improve signal.

Table 4 Signal-to-noise in various SRMs obtained by analysis of a 1200 pg/mL glucagon and GLP1 plasma extract, analysed using *m*-NBA or the original formic acid modified mobile phases.

Glucagon transition	S/N		GLP-1 transition	S/N	
	0.2% FA	0.05% MNBA 0.01% FA		0.2% FA	0.05% MNBA 0.01% FA
581.1/575.3	N/A	31.1	550.5/547.5	N/A	18.0
581.1/578.2	N/A	26.9	550.5/601.5	N/A	14.5
697.5/639.8	17.6	15.4	550.5/639.3	N/A	10.8
			660.1/656.7	12.0	3.4
			660.1/751.5	7.0	8.1
			825.1/458.2	N/A	N/A
			825.1/946.4	4.1	N/A
			1099.5/1093.7	N/A	N/A

Mobile phases are A= ACN and B=water (or water/ACN 95/5 for *m*-NBA modification), modified as shown in the table. N/A- not applicable, as no peak was observed

5.4 Comparison of SIL internal standard response of a surrogate matrix and extracted plasma sample

Figure S10 SIL gluagon internal standard response in a) a non-extracted surrogate matrix sample, and b) an extracted plasma sample

6 Establishment Procedures (Supplemental Tables)

Table 5 Precision and accuracy of surrogate matrix and extracted plasma glucagon samples analysed using LC-MS/MS with either the original formic acid modified mobile phases, or *m*-NBA modified mobile phases

Sample ID	Surrogate/ Plasma	Nominal concentration (pg/mL)	0.2% FA Phases Glucagon (697.5/694.0)		0.05% <i>m</i> -NBA 0.01% FA phases Glucagon (581.5/578.5)	
			%RE	%CV	%RE	%CV
QC LLOQ	Surrogate	15	110.3	11.9	96.9	15.4
QC LOW	Surrogate	45	113.8	6.2	110.9	11.8
(QC LLOQ)*	Plasma	25 ¹	92.6	7.0	90.6	10.0
(QC LOW)*	Plasma	75 ¹	92.4	11.3	99.6	10.9
QC MED	Plasma	200 ¹	91.4	8.7	100.0	9.7
QC MED (Lith Hep)	Plasma	200 ²	92.6	9.0	119.0	9.3
QC MED (inter matrix)	Plasma	200 ³	106.3	4.5	111.7	5.3
QC Dilution	Plasma	1750 ⁴	84.9	6.7	78.1	8.3
QC HIGH	Plasma	1750 ¹	98.0	10.7	106.1	12.2

Green statistics are within 20% RE or <20% CV for glucagon, or within 25% RE or <25 % CV for GLP-1

* = QC level created for GLP-1 analysis, but additionally monitored for glucagon to improve assay characterisation

1 = 31.9 pg/mL / 37.5 pg/mL endogenous glucagon measured using formic acid / *m*-NBA phases, and added to the spiked concentration

2 = No endogenous glucagon detected in lithium heparin plasma used

3= Adjusted for endogenous glucagon concentration determined in each individual matrix

4= No endogenous glucagon detected in matrix used for dilution QCs

Table 6 Glucagon matrix effects determined using LC-MS/MS with either the original formic acid modified mobile phases, or *m*-NBA modified mobile phases

Matrix ID	0.2% FA Phases Glucagon (697.5/694.0)			0.05% <i>m</i> -NBA 0.01% FA phases Glucagon (581.5/578.5)		
	Analyte ME	IS ME	ME Ratio	Analyte ME	IS ME	ME Ratio
F1	1.58	1.48	1.07	1.09	0.985	1.10
F2	1.36	1.43	0.951	1.01	1.05	0.961
F4	1.33	1.34	0.992	0.924	0.899	1.03
F5	1.30	1.47	0.882	1.16	1.19	0.979
M1	1.33	1.21	1.10	0.922	0.792	1.16
M4	1.32	1.20	1.09	0.656	0.591	1.11
Mean	1.37	1.36	1.01	0.960	0.918	1.06
SD	0.104	0.127	0.0881	0.176	0.209	0.0806
%CV	7.615	9.34	8.7	18.354	22.82	7.62

ME (Matrix Effect) = Adjusted Peak Area (Plasma)/ Peak Area (surrogate matrix)

Adjusted Area= Peak area of plasma ME sample- Peak area of endogenous plasma sample

ME ratio= Analyte ME/ IS ME

Green statistics have ME Ratios within 20% of 1, or %CV≤20%

224 **Table 7 Precision and accuracy of extracted plasma GLP-1 samples analysed using LC-MS/MS with formic acid (top) or *m*-NBA (bottom) modified**
225 **mobile phases with either the analogue or SIL glucagon internal standard.** Green statistics are within 25% RE or <25%CV. Red statistics are outside
226 these criteria. NA –not analysed, as 660.1/751.5 was determined to be optimal before these experiments were conducted.

Sample ID	Surrogate/ Plasma	Nominal concentration (pg/mL)	0.2% FA Phases GLP-1 (660.1/656.7)						0.2% FA Phases GLP-1 (660.1/751.5)					
			No IS		Heavy Glucagon		Des-7-Thr- Glucagon		No IS		Heavy Glucagon		Des-7-Thr- Glucagon	
			%RE	%CV	%RE	%CV	%RE	%CV	%RE	%CV	%RE	%CV	%RE	%CV
QC LLOQ	Plasma	25	102.2	18.2	88.0	11.6	86.9	16.7	131.4	20.1	118.6	20.0	112.7	24.6
QC LOW	Plasma	75	102.6	13.1	113.1	13.6	113.8	11.7	103.7	22.4	110.2	11.3	107.4	13.6
QC MED	Plasma	200	107.9	16.6	107.4	7.7	109.6	8.4	108.6	10.7	107.3	2.0	107.2	6.7
QC MED (Lith Hep)	Plasma	200	NA	NA	NA	NA	NA	NA	72.9	15.6	81.9	12.6	75.2	12.9
QC MED (inter matrix)	Plasma	200	NA	NA	NA	NA	NA	NA	149.1	15.4	134.5	16.0	122.9	16.3
QC DIL	Plasma	10,000	73.1	9.6	72.1	10.5	74.7	11.8	76.6	6.7	75.7	8.8	76.8	10.0
QC HIGH	Plasma	1750	99.7	15.6	111.6	2.7	113.9	5.8	102.9	15.5	113.8	3.6	113.8	6.7

Sample ID	Surrogate/ Plasma	Nominal concentration (pg/mL)	0.05% <i>m</i> -NBA 0.01% FA phases (550.5/601.5)						0.05% <i>m</i> -NBA 0.01% FA phases (550.5/693.3)					
			No IS		Heavy Glucagon		Des-7-Thr- Glucagon		No IS		Heavy Glucagon		Des-7-Thr- Glucagon	
			%RE	%CV	%RE	%CV	%RE	%CV	%RE	%CV	%RE	%CV	%RE	%CV
QC LLOQ	Plasma	25	88.0	25.6	75.1	36.6	75.1	33.8	75.0	43.2	57.5	39.8	63.4	33.0
QC LOW	Plasma	75	83.9	17.3	107.0	19.6	103.9	15.8	71.0	15.5	104.4	25.3	96.0	22.8
QC MED	Plasma	200	80.2	14.1	94.8	8.8	94.4	7.2	75.0	13.6	91.4	11.5	88.3	12.8
QC MED (Lith Hep)	Plasma	200	NA	NA	NA	NA	59.9	19.4	NA	NA	NA	NA	NA	NA
QC MED (inter matrix)	Plasma	200	NA	NA	NA	NA	177.1	37.5	NA	NA	NA	NA	NA	NA
QC DIL	Plasma	10,000	63.8	12.4	75.8	7.4	74.6	8.3	63.6	13.7	83.1	7.2	76.6	7.9
QC HIGH	Plasma	1750	80.1	12.8	99.6	16.8	107.4	12.2	78.7	13.6	103.2	18.4	106.8	14.1

Table 8 GLP-1 matrix effects determined using LC-MS/MS with the original formic acid modified mobile phase and the analogue glucagon internal standard. See footnotes on Table 6.

Matrix ID	0.2% FA Phases GLP-1 (660.1/751.5) with analogue glucagon IS		
	Analyte ME	IS ME	Ratio
F1	1.14	1.41	0.808
F2	0.869	1.33	0.653
F4	0.876	1.40	0.628
F5	0.735	1.34	0.548
M1	1.02	1.24	0.827
M4	0.765	1.19	0.641
Mean	0.901	1.32	0.684
SD	0.154	0.0860	0.110
%CV	17.1	6.5	16.1

Table 9 Inter-batch reproducibility of endogenous glucagon determination.

Replicate	Measured Endogenous Glucagon Concentration (pg/mL)	
	Initial Analysis	Reanalysis after 42 Days storage at -80°C
1	29.7	34.0
2	31.1	36.5
3	33.8	31.7
4	32.7	35.7
5	34.2	33.8
6	29.9	27.9
Mean	31.9	33.3
SD	1.95	3.11
CV	6.1	9.4
% Difference	4.3	

Formic acid modified mobile phases used.
 $\% \text{ Difference} = 100 * ((\text{Reanalysis concentration} - \text{Initial concentration}) / \text{Initial concentration})$

237

Table 10 Performance of QCs in stabilised plasma vs non-stabilised plasma.

Anticoagulant	Stabiliser	Freeze-thaw cycles	% Non-stabilised QC concentration	
			Glucagon	GLP-1
EDTA	None	0	113.3	77.9
EDTA		0	94.8	99.7
EDTA	D	0	102.9	104.9
EDTA	AD	0	109.1	107.0
EDTA	AD	1	101.1	86.2
EDTA	AD	2	109.5	95.0
Lithium Heparin	None	0	81.7	84.1
Lithium Heparin	A	0	98.1	97.1
Lithium Heparin	D	0	91.1	105.9
Lithium Heparin	AD	0	98.0	113.7

238

Formic acid modified mobile phases used

239

200 pg/mL of glucagon/GLP-1 was spiked into a variety of matrices, as shown, and compared to the performance of QCs spike in non-stabilised EDTA plasma.

240

241

Green statistics are within 20/25% (Glucagon/GLP) of non-stabilised QC response.

242

243

Table 11 Ability to reinject sample extracts after 42 days storage at 4°C

Surrogate/ Plasma Matrix	Nominal Concentration (pg/mL)	Glucagon		GLP-1	
		%RE	%CV	%RE	%CV
Surrogate	15	107.1	11.6	N/A	N/A
Surrogate	25	106.0	11.3	N/A	N/A
Surrogate	45	111.3	7.8	N/A	N/A
Plasma	25#	86.5	7.6	116.6	7.4
Plasma	75#	87.7	10.1	99.3	24.8
Plasma	200#	89.6	10.0	98.4	7.3
Plasma	1750#	99.0	13.6	104.7	22.2

244

Formic acid modified mobile phases used.

245

#31.4 pg/mL endogenous concentration added to glucagon QC concentration

246

Green statistics are within 20% RE or <20%CV (Glucagon) or 25% RE or <25%CV (GLP-1).

247 **7 Cross validation against established immunoassays using physiological samples**

248

249

250 **Figure S11 Comparison of endogenous glucagon concentrations determined by LC-MS/MS, Milliplex and HTRF immunoassays.**
251 Samples were analysed from 5 volunteers (1-5), which contained either EDTA (E) or Lithium Heparin (L) anticoagulant in addition to various stabilisers (not
252 shown). Different stabiliser combinations were used in samples of the same Sample Type, e.g. the four 4E samples in Figure 11a are all unique. a) All
253 concentrations determined. b) Concentrations above the LLOQ only shown ($\geq 10.0/13.0/12.0$ pg/mL for LC-MS/MS / Milliplex / HTRF respectively).

254 **8 References**

255 [1] Howard JW, Kay RG, Tan T, et al. Development of a high throughput UHPLC-MS/MS (SRM)
256 method for the quantitation of endogenous glucagon from human plasma. *Bioanalysis*. 6(24),
257 3295–3309 (2014).

258 [2] Iavarone AT, Williams ER. Mechanism of charging and supercharging molecules in
259 electrospray ionization. *J. Am. Chem. Soc.* 125(8), 2319–27 (2003).

260