This item was submitted to [Loughborough's Research Repository](#) by the author.
Items in Figshare are protected by copyright, with all rights reserved, unless otherwise indicated.

Potential of microneedle-assisted micro-particle delivery by gene guns: a review

PLEASE CITE THE PUBLISHED VERSION

<http://dx.doi.org/10.3109/10717544.2013.864345>

PUBLISHER

© Informa Healthcare USA, Inc.

VERSION

AM (Accepted Manuscript)

LICENCE

CC BY-NC-ND 4.0

REPOSITORY RECORD

Zhang, Dongwei, Diganta Bhusan Das, and Chris D. Rielly. 2014. "Potential of Microneedle-assisted Micro-particle Delivery by Gene Guns: A Review". figshare. <https://hdl.handle.net/2134/14349>.

This item was submitted to Loughborough's Institutional Repository (<https://dspace.lboro.ac.uk/>) by the author and is made available under the following Creative Commons Licence conditions.

For the full text of this licence, please go to:
<http://creativecommons.org/licenses/by-nc-nd/2.5/>

Potential of microneedle assisted micro-particle delivery by gene guns: A review

Dongwei Zhang, Diganta B Das*, Chris D Rielly

Department of Chemical Engineering, Loughborough University, Loughborough
LE113TU, Leicestershire. UK

Running title: Microneedle assisted micro-particle delivery

Accepted in 'Drug Delivery'

6 November 2013

*Corresponding Author (Email: d.b.das@lboro.ac.uk)

Potential of microneedle assisted micro-particle delivery by gene guns:

A review

Dongwei Zhang, Diganta B Das*, Chris D Rielly

Department of Chemical Engineering, Loughborough University, Loughborough LE113TU,
Leicestershire. UK

(*Corresponding Author; Email: d.b.das@lboro.ac.uk)

Abstact

Context

Gene guns have been used to deliver deoxyribonucleic acid (DNA) loaded micro-particle and breach the muscle tissue to target cells of interest to achieve gene transfection.

Objective

This paper aims to discuss the potential of microneedle (MN) assisted micro-particle delivery from gene guns, with a view to reducing tissue damage.

Method

Using a range of sources, the main gene guns for micro-particle delivery are reviewed along with the primary features of their technology, e.g., their design configurations, the material selection of the micro-particle, the driving gas type and pressure. Depending on the gene gun system, the achieved penetration depths in the skin are discussed as a function of the gas pressure, the type of the gene gun system, and particle size, velocity and density. The concept of MN-assisted micro-particles delivery which consists of three stages (namely, acceleration, separation and decoration stage) is discussed. In this method, solid MNs are inserted into the skin to penetrate the epidermis/dermis layer and create holes for particle injection. Several designs of MN array are discussed and the insertion mechanism is explored, as it determines the feasibility of the MN based system for particle transfer.

Results

The review suggests that one of the problems of gene guns is that they need high operating pressures, which may result in direct or indirect tissue/cells damage. MNs seem to be a

promising method which if combined with the gene guns may reduce the operating pressures for these devices and reduce tissue/cell damages.

Conclusions

There is sufficient potential for MN-assisted particle delivery systems.

Key words: gene gun, microneedle, micro-particle, skin, penetration depth

1. Introduction

A gene gun, involving “particle bombardment” or “biolistic delivery” is a particle accelerator, which can deliver deoxyribonucleic acid (DNA) loaded micro-particles at sufficiently high velocities to breach the surface of target tissue and to penetrate to a depth to achieve gene transfection in the cells. A gene gun is considered to be a unique concept which was first used to deliver genetic materials into plant cells (Klein et al., 1987; Klein et al., 1992; Sanford et al., 1993; Svarovsky, 2008; Huang et al., 2011; O'Brien et al., 2011; Manjila et al., 2013). The technique has been used to transfer DNA-coated micro-particles to achieve gene transfection into many types of cells and organs (e.g., Meacham et al., 2013; Bennett et al., 1999), for example, mammals (Cao et al., 2013; Ettinger et al., 2012; Da'dara et al., 2002; Kuriyama et al., 2000; Sakai et al., 2000; Williams et al., 1991), plants (Kuriakose et al., 2012; Zuraida et al., 2010; Klein et al., 1992), artificially cultured cells (O'Brien and Lummis, 2011, 2006), fungi (Gu et al., 2011; Armaleo et al., 1990) and bacteria (Nagata et al., 2010; Smith et al., 1992). A number of commercial gene guns have been manufactured and used for *in vivo* gene transfection in plants, living animals, cultured cells, e.g., the PowderJect system developed originally by the University of Oxford (UK) (Bellhouse et al., 2006), the Helios gene gun by Bio-rad, Hercules, CA (USA) (Belyantseva, 2009), and the SJ-500 portable gene gun by Biotech instrument, New Jersey (USA). The Helios and portable gene guns are shown in Figure 1. In a research context, Da'dara et al. (2002) have used a Helios gene gun to delivery Sm23-pcDNA (an integral membrane protein) to mice to evaluate the immunogenicity of the protective efficacy of the DNA vaccination. Ahlen et al. (2013) have used a Helios gene gun to explore a method for monitoring hepatitis B and C viruses (HBV/HCV) specific to immune responses in mouse.

Loading the micro-particles with DNA (for example onto gold particles) is an important step in using this technology, as discussed by Zhao et al. (2012); O'Brien and Lummis (2006, 2011), Rao (2010), Satish, (2009), Svarovsky et al. (2008), and Thomas et al. (2001). However, a detailed discussion on these issues is outside the scope of this paper.

Such guns have been proposed for gene delivery in the treatment of some of the major diseases such as cancer (Nguyen-Hoai et al., 2012; Aravindaram et al., 2009; Han et al., 2002; Chen et al., 2002); these researchers have fired gene loaded micro-particles into mammals or cultured cells instead of humans to study gene expression and cell damage. For example, Yoshida et al. (1997) used a gene gun to deliver DNA coated Au micro-particles into the liver of living rat at 250 psi (~17.2 bar) pressure, which resulted in a good gene expression but caused a number of cell deaths. The work also showed that cell damage was not obtained below 150 psi of pressure. Sato et al. (2000) have used various types of gene guns to transfer genes into live rodent brain tissue, which confirmed gene expression, but with mechanical damage to cells. Uchida et al. (2009) have fired plasmid DNA into cultured mammalian cells (e.g., human embryonic kidney (HEK293) cell, human breast adenocarcinoma (MCF7) cell) using a gene gun, which showed that transfection is achieved in the cells, but the cell damage occurs at operating pressures greater than 200 psi (~13.8 bar). O'Brien and Lummis (2011) have cultured HEK293 cells as a target for the biolistic transfection using a gene gun. Their work showed that nano-particles gave a similar performance to micro-particles for biolistic transfection, but created less cell damage.

The above works show that cell damage may be a problem for biolistic micro-particle delivery, due to significant gas and particle impactions on the tissue. This also means that while gene guns have been reported to be successful in many instances, they may have some disadvantages. Generally, for most gene gun systems, the maximum penetration of the micro-particle can breach the *stratum corneum* and end inside the epidermis layer of the tissue (Yager et al., 2013; Liu, 2006; Quinlan et al., 2001; Bennett et al., 1999). The epidermis layer is normally considered to be a target site for gene delivery due to its accessibility (Soliman et al., 2011b; Liu, 2006; Quinlan et al., 2001; Trainer and Alexander, 1997). However, a high

pressure is required for most gene guns, which might damage the tissue after the impact of the pressurized gas on the skin.

Figure 1: Left: The commercial hand-held gene gun of SJ-500 (Biopex, 2012), Right: the commercial Helios gene gun (Bio-rad, 2012)

Cell damage can be reduced by decreasing the particle size and/or operating pressure, as both reduce the impact forces. In order to satisfy these conditions and, possibly, increase the penetration depth of the micro-particle in the target, an innovative concept of combining MNs with a particle delivery system, namely MN-assisted micro-particle delivery, has been proposed recently by Zhang et al. (2013). The MNs are discussed in section 3 of this paper. The operating principle of the delivery in the approach of Zhang et al. (2013) is that a pellet of micro-particles is loaded on a ground slide which is accelerated by a pressurized gas to a sufficient velocity. The pellet is released after the ground slide reaches the end of a barrel in the system, and separates into micro-particles of a narrow size distribution by impactation on an open mesh; these separated micro-particles then penetrate into the desired target. The resistance of the target/skin to the penetration of the micro-particles is reduced by using MN to create a number of holes through which the micro-particles can enter, without the need for very high gas pressures.

MN assisted micro-particle delivery seems to be a promising approach for dry particulate delivery, but the potential of this method needs to be discussed thoroughly as it is still at an early stage of development. To this end, the main types of gene gun systems for micro-particle delivery and their operating principles are reviewed. The micro-particle material and size for

these gene guns are discussed, and the range of gas pressures and particle velocities are considered.

In this paper, MN arrays are discussed to understand how various geometries affect penetration of the target material and formation of holes, which remain after the needles are removed. This discussion provides relevant background knowledge and a sound foundation for further improvement of gene gun systems, with the help of MN assisted micro-particle delivery to achieve similar levels of penetration depth, but with less cell damage. To further understand the advantage of MN assisted micro-particle delivery, a comparison with other physical cell targeting approaches is discussed. Finally, some simplified models of the micro-particle delivery in the skin are described, which will be useful in predicting the penetration depths achieved by MN assisted gene gun delivery.

2. Main gene guns

2.1 Configurations and operating conditions

Based on the principles of the operation process and driving forces, we define that the gene guns can be divided into three types, namely: powder gene gun, high-voltage electric gene gun and gas gene gun.

A powder gene gun is the original device which uses an ignition gunpowder as a driving force to promote the movement of bullets of macro-projectiles, thus accelerating the micro-particles which are loaded onto the bullet (Huang et al., 2011; O'Brien and Lummis, 2007; Zhou, 1995; Klein et al., 1987; Sanford et al., 1987). An electric gene gun uses a high voltage (HV) to vaporize water droplets; the expanding gas is used to achieve the acceleration of micro-particles (Christou et al., 1990). A gas gene gun releases a high pressure gas to accelerate the micro-particles (or micro-particles loaded ground slide) to a sufficient velocity to breach the barrier of the target (Zhang et al., 2013; Soliman et al., 2011b; Liu, 2007; Kendall et al., 2002; Zhou, 2000).

Powder gene guns were originally developed to deliver genetic material coated micro-particles into plant cells (Klein et al., 1987; Klein et al., 1992; Svarovsky, 2008; O'Brien et al., 2011; Manjila et al., 2013). As shown in Figure 2, an explosion of gun powder accelerates

DNA-coated micro-particles attached onto the front surface of a macro projectile. Klein et al. used tungsten micro-particles of 4 μm average diameter, which could be accelerated to 400 m/s. However, the explosion of the gun powder is noisy and may cause cell/tissue damage.

Figure 2: A schematic sketch of the powder gun (redrawn from Klein et al. (1987))

Zhou (1995) improved the design of Klein et al. (1987) and patented a powder gun (Figure 3) which is composed of a gun body, a gun barrel, a ground slide, a baffle plate and a baffle plate fixed pipe. An ignition of gunpowder, caused the bullet to accelerate and hit a ground slide, which then impacts on a baffle plate. Thus, the micro-particles gain an initial velocity to leave the ground slide, pass through the central hole of the baffle plate and penetrate into the tissue. The advantage of this powder gene gun is that the use of ground slide can reduce explosion damage to tissue. The base is designed as a shock absorber to reduce the recoil force, using a spring attached to the firing pin.

Figure 3: The layout of powder gene gun (redrawn from Zhou (1995))

Christou et al. (1990) have used a type of electronic gene gun to deliver DNA-coated gold micro-particles into soybean seeds; a schematic sketch of the high-voltage electric gene gun is shown in Figure 4. Yang et al. (1990) have applied an electronic gene gun to transfer genes

into the muscle, tissue and liver of a mouse, using 1-3 μm diameter golden particles. The disadvantage of their gun is that a high voltage up to 18 kV is required to deliver a micro-particle penetration depth of 125 μm . Recently, Ikemoto et al. (2012) used a type of high-voltage electric gene gun, namely, an electrospray device to accelerate liquid droplets in a high voltage (12 kV) to collide with DNA coated micro-particles and deliver them into living cells.

Gas gene guns can be classified into two different delivery methods. The first uses a high pressure gas to push the micro-particle coated ground slide to achieve the goal of particle acceleration. The disadvantage is that a very high operating pressure is required to achieve the desired micro-particle impact velocity. However, the advantage is that the ground slide prevents the released gas from impacting on tissue. The second method mainly uses a high pressure gas (e.g., N_2 , He) to directly push the DNA-coated micro-particles to a sufficient velocity to breach the *stratum corneum*, penetrate into the epidermis/dermis layer of the skin or deeper tissues. The disadvantage is that very high gas pressure may damage the tissue and the muscles. The advantage is that the micro-particle can reach a desired velocity under lower operation pressure as the micro-particles have very small mass.

Figure 4: The schematic sketch of the high-voltage electric gene gun (redrawn from Christou et al. (1990))

Using the principles of the first delivery method, Zhou (2000) has invented a kind of high-pressure gene gun which consists of a casing, the compressed gas inlet pipe, emitting cavity and bombarding cavity. The schematic diagram of this gun is shown in Figure 5 which shows both the states of the gun before and after operation. The emitting cavity consists of an air storage house, membrane, bullet, baffle plate and sample holder. The membrane will be

ruptured when the air storage house reaches a certain pressure. High pressure gas is able to
 accelerate the bullet to the baffle plate. Since the bullet is blocked by the baffle plate, the
 coated particles will leave the surface of the bullet, go through the centre hole of the baffle
 plate and launch into the sample, thus completing the gene injection. In addition, the gas will
 be released from the vent hole. It has been claimed that this instrument has good stability, high
 efficiency, does not produce impurities, and particles can also attain a higher initial speed.
 However, it is generally applied for plant tissues.

Mitchell et al. (2003) have conducted many experimental and numerical studies on the light
 gas gun (LGG), as shown schematically in Figure 6. The LGG uses a high pressure helium
 gas to drive the micro-particle coated ground slide to a certain velocity. In this case, the
 polystyrene particles of 99 μm diameter have been used and the impact velocities of the
 particles have been shown to reach 170, 250 and 330 m/s under 20, 40 and 60 bar operation
 pressures.

Figure 5: The schematic diagram and principle of high-pressure gas gene gun (redrawn from
 Zhou (2000))

Figure 6: The schematic sketch of light gas gun (redrawn from Mitchell et al. (2003))

Williams et al. (1991) have studied a helium-driven gene gun which is somewhat similar to the design of Zhou's (2000), as shown in Figure 7. In this case, a membrane in the system breaks after the gas pressure reaches a certain value. The micro-projectiles are accelerated by the helium gas and separated well by a stopping screen. In addition, the large particle is blocked by a screen to avoid tissue damage. In this case, golden particles of 1 - 3 μm and 2 - 5 μm diameters have been fired and penetration depths of 150 μm and 200 μm have been obtained in a mouse liver, under 1300 psi operating pressure.

Figure 7: The schematic sketch of helium-driven gene gun (redrawn from Williams et al. (1991))

Kendall (2002) has reported a contoured shock tube which is shown in Figure 8. In this case, the compressed gas will pressurize the membrane, and the micro-particle will be accelerated

by a shock wave as the gas pressure rises to the point where the membrane ruptures. Liu et al. (2004a) and Liu (2006) have described an injection device, namely PowderJect (Figure 9), which uses helium gas as the source of momentum. A trigger actuates a mechanism to release helium gas, which expanded to accelerate the micro-particles to a sufficient momentum to pierce the outer layer of the target and into the cells of interest.

Figure 8: The schematic diagram of contoured shock tube (redrawn from Kendall (2002))

Figure 9: A schematic diagram of the PowderJect system (redrawn from Liu (2006))

Zhang et al. (2007) have introduced the principle of the Helios gene gun which contains acceleration, separation and deceleration stages. The process of micro-particle delivery is shown in Figure 10. This gene gun uses helium gas to accelerate DNA-coated micro-carriers which are separated by a stopping screen. The separated micro-particles exit from the gene gun at high speed, penetrate the tissue to the targeting area, enter into the cell and hit the nucleus membrane.

Figure 10: Gene transfer stages of the biolistic gene gun (redrawn from Zhang et al. (2007)); L_a , L_s and L_d are the distance of acceleration, separation and deceleration stages.

In addition, Zhou (2007) has created a special liquid gene sprayer (Figure 11) which is different from the above three gene gun types. It has been used to deliver liquid-form medication into the human body and consists of an interconnected casing and magazine. The magazine comprises of a hollow cylinder and sliding piston rod set; emission holes are arranged at the front end of the cylinder and the front end of the casing is fixed to rear end opening. An energy-storage driving mechanism consists of a spring and impeller and set in the internal cavity of the casing. The impeller moves backward to press the spring to store the energy, and pushes forward by a driving force from the released spring. This device uses a small volume energy storage device to inject the required amount of biological gene, and does not need any separate air supply equipment; it is easy to handle and carry. It uses the liquid as the DNA particle's carrier and hence golden particles are not required in this device. The DNA particles will be suspended in the liquid. Before using this device, the head of the hollow cylinder is inserted into the DNA particle coated liquid, and the sliding piston rod is pulled out to extract the liquid and load the gun. Then the spring is compressed very tightly, storing energy (see Figure 11a), which is released to accelerate the impeller and sliding piston rod; then the DNA particle coated with liquid are pushed out from the emission hole and penetrate into the tissue of plant or human to the target cells of interest. The state of the liquid gene gun after injection is shown in Figure 11b.

Figure 11: The liquid gene sprayer: A: The liquid gene gun at energy storage state, B: The state of liquid gene gun after injection (redrawn from Zhou (2007))

Figure 12: The schematic of the laser plasma jet (redrawn from Menezes et al. (2012))

Recently, Menezes et al. (2012) have designed an advanced laser plasma jet (see Figure 12) to deliver DNA-coated micro-particle. The operating principle is that a laser beam is fired and ablates a thin aluminum foil, using lenses for focusing. The laser ablation is confined by the BK7 glass overlay to improve performance. Thus, it causes the foil to evaporate into an ionized vapor and the sudden blow-off causes a shock wave to breach the foil to accelerate DNA

coated micro-particles. The device provides micro-particles impact velocities of up to 1100 m/s, which is faster than other gen guns (e.g. CST, LLG). However, this technique is costly due to the use of laser ablation.

2.2 Micro-particle materials and size

The materials and size of the micro-particles which are used in gene guns have significant importance on the operation of the system, e.g. by determining the routes and extent of particle penetration into tissue. In general, the routes of the micro-particle penetration in the tissue are normally divided into two types, which are the extracellular and intercellular routes (Bryan et al., 2013; Soliman, 2011; Mitchell et al., 2003). As presented in Figure 13(a-b), the penetration routes of the micro-particles in tissue depend on the particle size. The extracellular route is followed for large particle delivery, e.g., for epidermal powder immunization (Soliman, 2011; Hardy et al., 2005). An illustration of the range of particle material and sizes for the relevant gene gun systems is listed in Table 1. It shows that the extracellular route is normally followed for the less dense materials (e.g., stainless steel, polystyrene and glass) with diameters ranging from 15 to 99 μm . Hardy et al. (2005) have reported that particle diameters ranging from 25 to 100 μm are expected to follow the extracellular route, as their momentum is insufficient to breach the target barrier of target, due to the combination of relatively low density and small size. It has been recommended that stainless steel or polymer micro-particles should be used for extracellular routes, due to their biocompatibility and low cost (Soliman, 2011; Sung et al., 2011; Singh and Dahotre, 2007; Binyamin et al., 2006; Disegi et al., 2000).

Figure 13: Schematic diagram of the extracellular and intercellular failure mechanisms (a): extracellular failure mechanism for large particles (b): intercellular failure mechanism for small particles (redrawn from Mitchell et al. (2003)).

In contrast, the intercellular route (e.g. for DNA immunization), uses smaller size, but much more dense gold or tungsten micro-particles (Soliman, 2011; Mitchell et al., 2003). In order to deliver DNA into cells effectively, dense materials are preferred which are prepared into micro-particles of diameters ranging from 0.6 to 6 μm (Soliman, 2011; Rao, 2010; Hardy et al., 2005) which are also smaller than the cell diameters. High-speed micro-particles breach the skin and may penetrate through the individual cell membranes. It is well known that the most recommended material of these micro-particles for gene gun system is gold due to its high density, low toxicity and lack of chemical reactivity (Valenstein, 2012; Rosi et al., 2006; Macklin, 2000). However, tungsten micro-particles have also been used as micro carriers in gene gun systems, due to their lower cost. Tungsten particles have some disadvantages for genetic transformation, such as non-biocompatibility and toxicity (Bastian et al., 2009; Yoshimisu et al., 2009; Russell et al., 1992). Recently, Hou et al. (2013) have used titanium dioxide (TiO_2) for biolistic micro-particle delivery due to their biocompatibility (Singh and Dahotre, 2007) and low density (2 g/cm^3) which may reduce the cell damage after particle impactation. The particle impactation may cause cell damage (O'Brien and Lummis, 2011; Sato et al., 2000), which is a significant area for gene gun research.

2.3 Gas pressure and particle velocity

Up to now, gene gun systems have been widely used to deliver DNA loaded micro-particles into cells for research of DNA transfection, e.g., CST (Rasel et al., 2013; Liu, 2008; Truong et

al., 2006; Liu et al., 2006; Kendall, 2002) and Helios gene gun (O'Brien and Lummis, 2011; Belyantseva, 2009). Helium and compressed air gases are often used as driving forces to accelerate micro-particles for gene gun system. Especially, helium gas is recommended for most gene gun systems due to its non-toxic, low density, lack of chemical inactivity and high compressibility factor (Marrion et al., 2005), which allow the particles to reach higher velocities (Tekeuchi et al., 1992). Compressed air is often used as a substitute for helium due to its lower cost.

The gas pressure is a major factor which should be considered for gas gene gun systems. It directly affects the velocity of the micro-particles, e.g., Liu et al. (2008) demonstrated that 1.8 μm diameter gold micro-particles can reach a velocity of 580, 650, 685 and 710 at 3, 4, 5 and 6 MPa, respectively. In addition, the velocity is also related to the micro-particle size and density: particle velocity is increased from an increase in operating pressure and a decrease in particle size.

334 **Table 1:** Illustration of the particle material and size for the relevant gene gun systems

Type of gene gun	Material of particles	Average diameter of micro-particles (µm)	Reference
Powder particle gun (PPG)	Tungsten	4	Klein et al. (1987)
Helium-driven apparatus (HDA)	Tungsten	3.9	Williams et al. (1991)
	Gold	1-3, 3-5	
Conical nozzle (CN)	Polymeric	4.7, 15.5 and 26.1	Quinlan et al. (2001)
Converging-diverging nozzle (CDN)	Polystyrene	4.7	Kendall et al. (2004a)
Pneumatic gun (PG)	Gold	0.47 ± 0.15, 1.1 ± 0.1	Rinberg et al. (2005)
	Silicon	2 – 18	Zilony et al. (2013)
	Gold	1.6	
Light gas gun (LGG)	Stainless steel	25	Mitchell et al. (2003)
Contoured shock tube (CST)/LGG	Polystyrene	15.5, 25.2, 48 and 99	
CST	Gold	3.03	Truong et al. (2006)
	Polystyrene	15 and 48	
	Glass	46	
	Gold	2.7 and 3.5	
	Polystyrene	39 ± 1	Liu et al. (2006)
Helios gene gun	Gold	0.6, 1.0 and 1.6	Uchida et al. (2009)
		1	O'Brien and Lummis (2011); Kuriakose et al. (2012); Cao et al. (2013)
Biolistic PDS/1000 Helium system (BioRad, USA)	Gold	0.6, 1.0 and 1.6	Zuraida et al. (2010)
BioWare low pressure gene gun	Gold	1	Yen and Lai (2013)

Along with the development of gene gun technology, the achievable particle velocity and penetration depth in the target vary between gene gun systems. For example, Quinlan et al. (2001) have applied a conical nozzle to accelerate polymeric micro-particles of 4.7, 15.5 and 26.1 μm diameters, to reach velocities of 350, 460 and 465 m/s at 60 bar pressure, respectively. A contoured nozzle has been tested by Quinlan et al. (2001) who employed 60 bar to accelerate polymeric micro-particles of 4.7 μm diameter to a velocities of about 1000 m/s, but 26.1 μm diameter of particles only reached 740 to 810 m/s. The velocity of polymeric micro-particles of 15.5 μm diameter only reaches 330 m/s at 60 bar for the LGG system (Mitchell et al., 2003). Kis et al. (2011) have concluded that the particle velocity reached about 700 m/s for CST (Liu et al., 2006) and range from 200 to 800 m/s for converging-diverging nozzles (Liu et al., 2004b; Kendall, 2002; Quinlan et al., 2001). Recently, Menezes et al. (2012) have operated an advanced laser plasma jet to drive gold micro-particles of 1 μm diameter to achieve an average velocity of 1100 m/s within a distance of only 10 mm. The particular achievements (e.g., particle velocity, operating pressure and penetration depth) for various gene gun systems are listed in Table 2.

From the above studies, it can be concluded that the operating pressures for gene gun systems generally vary from 20 to 60 bars. Xia et al. (2011) have suggested that the pressure should be held below about 13 bar to minimise damage from the impaction of pressurized gas on soft tissue. Uchida et al. (2009) achieved gene transfection in cells but noted damage if the operating pressure is over 13 bar. These reports demonstrate that cell death is unavoidable for many of the current generation of gene gun systems. Mitchell et al. (2003) show that golden micro-particles of 3.03 μm diameter employed at 60 bar pressure in the CST only reach a maximum penetration depth of 60 μm in the canine buccal mucosa. Normally, the viable epidermis layer of skin is the target area for gene gun systems. Mitchell et al.'s (2003) results show that micro-particles require a still higher velocity to penetrate through the *stratum corneum*, which means higher operating pressures are necessary. However, Mitchell et al. (2003) also show stainless steel micro-particles of 25 μm diameter can achieve 124 μm penetration depths in the canine buccal mucosa at 20 bar pressure (see Table 2). But O'Brien and Lummis (2011) show that cultured cells are damaged by gold micro-particle of 1 μm diameter, when operating at 3.4 bar pressure using a Helios gene gun. They indicate that

stainless steel micro-particles of 25 μm diameter impact on the tissue at a velocity of 170 m/s will damage the target tissue and cells. Thus, a new concept of applying MN to micro-particle delivery, which may reduce the cell damage, is discussed in section 4.2.

Table 2: Illustration of the materials and size of micro-particles used in gene gun systems

Gene gun	Material and average diameter of micro-particles	Operating pressure (bar)	Velocity at impact (m/s)	Target and maximum Penetration depth in target	Reference
PPG	Tungsten, 4 μm	N/A	430	Onion, 40 μm	Klein et al. (1987)
HAD	Gold, 1 - 3 μm	90	N/A	Mouse liver tissue, ~ 130 μm	Williams et al. (1991)
CN	Polymers, 4.7, 15.5 and 26.1 μm	60	350, 460 and 465	N/A	Quinlan et al. (2001)
LGG	Stainless steel, 25 μm	20	170	Canine buccal mucosa, 124 μm	Mitchell et al. (2003)
CST	Gold, 3.03 μm	60	550	Canine buccal mucosa, 60 μm	
CST	Gold, 1 $\pm$ 0.2 μm	40	580	Human skin, 66 μm	Kendall (2002)
CST	Polystyrene, 39 μm	60	570 $\pm$ 14.7	N/A	Liu et al. (2006)
Helios Gene Gun	Gold, 40 nm and 1 μm	5	N/A	Mouse ear tissue, 50 $\pm$ 11 μm and 31 $\pm$ 6 μm	O'Brien and Lummis (2011)
Laser plasma jet	Gold, 1 μm ; Tungsten, 1 μm	N/A	1100	N/A	Menezes et al. (2012)

3. MNs

3.1 Types and configurations of MNs

Henry et al. (1998) are widely regarded as the first to have developed a method of transdermal drug delivery using MNs, which has gradually developed for various applications of drug delivery. MN arrays are minimally invasive device that bypass the outer layer of skin, namely the *stratum corneum*, to achieve enhanced transdermal drug delivery (Olatunji et al., 2013; Donnelly et al., 2012; Prausnitz and Langer, 2008). MN are normally separated into two categories, namely, solid and hollow (Koelmans et al., 2013; Han et al., 2008; Qi et al., 2007). Each of these can be made of different materials and used for various functionalities, depending on their designs. The most common materials used for fabricating MN are metal, silicon, and polymer (Kim et al., 2012; Memon et al., 2011; Zhao et al., 2006). The primary metals used for MNs are stainless steel (Kim et al., 2012; Quan et al., 2010; Bal et al., 2008; Martanto et al., 2004) and titanium (Kim et al., 2010; Fernandez et al., 2009; Parker et al., 2007). Metal MNs have the advantages of low cost, tougher hardness, ease of penetration into the tissue and they are not easily broken in the tissue. Silicon and silica materials have better biocompatibility than metallic materials, but they are expensive (Chen et al., 2013, 2008). Furthermore, they break up more easily and fragments may be left inside the tissue after MN removal (Memon et al., 2011; Zhao et al., 2006). However, silicon is the first material used to fabricate MNs for pre-treatment of skin prior to patch application, e.g., Henry et al. (1998) have used solid conical silicon MNs (Figure 14a) with a height of 0.15 mm, inner diameter of 80 μm and tip diameter of 1 μm to increase skin permeability and to provide an effective delivery of drugs to diffuse through the skin. Mikszta et al. (2002) used a silicon MN array (see Figure 14b) for gene delivery in skin.

Polymeric materials are a cheap option which can exhibit biocompatibility and biodegradability, but the hardness is generally lower (Nayak and Das, 2013; Oh et al., 2008; Han et al., 2007; Park et al., 2005). Various types of polymer have been used for fabricating MNs, such as poly-glycolic acid (PGA) (Park et al., 2006, 2005) and polycarbonate (PC) (Han et al., 2007). Park et al. (2005) have fabricated a PGA solid MN array with needle length of 1500 μm , base diameter of 200 μm tapering to tip diameter of 20 μm . An example of solid polymer MN array is shown in Figure 15a. Recently, Donnelly et al. (2011) have used poly (vinyl) alcohol (PVA) to

404 fabricate solid MN arrays with a needle height of 600 μm , base width of 300 μm , and MN
405 interspacing of 300 μm . As can be seen from Figure 15 (b), MNs have been fabricated into
406 uniform conical-shaped needles.

408
409 Figure 14: (a). Solid conical MN arrays (Henry et al., 1998) (b) Silicon MN array used for gene
410 delivery in skin (Mikszta et al., 2002).

413
414 Figure 15: (a): Solid MN array made with the help of PGA (Park et al., 2005) (b): Solid MN
415 array made with the help of PVA (Donnelly et al., 2011)

Hollow MN arrays are normally used for fluid infusion of liquid drug and nanoparticles into the skin (Han et al., 2008). However, they are not widely used due to their cost and complex fabrication methods (Zhu et al., 2012). An example of the hollow MN array designed by Stoeber and Liepmann (2000) is shown in Figure 16a. It is a hollow conical MN array with a height of 200 μm and a channel diameter of 40 μm . McAllister et al. (2000) have improved Henry et al.'s (1998) design, by inserting a hole in the centre of each MN to fabricate a hollow silicon MN array.

Figure 16: (a) Hollow conical MN arrays on the right (Stoeber and Liepmann, 2000) (b) hollow silicon MN array with sloping side walls (Gardenier et al., 2003)

Subsequently, Gardenier et al. (2003) have designed a hollow silicon MN array with sloping side wall, which is shown in Figure 16b. The length of the needles varies between 150 and 350 μm , with a base diameter of 250 μm (measured at the widest section) and a maximum hole width of 70 μm . The centre of the hole is positioned 40 μm from the tip of the needle. Davis et al. (2004) have used a hollow metal MN array (Figure 17) with a tip diameter of 75 μm , base diameter of 300 μm , wall thicknesses of 5 μm and height of 500 μm to measure the required force for the insertion of MNs into the tissue. They reported that a force ranging between approximately 0.1 – 3 N is sufficient to penetrate a single MN into the tissue.

Figure 17: Hollow metal MNs (Davis et al., 2004)

Until now MNs have been developed as a minimally invasive means to deliver genes via the transdermal route (Tuan-Mahmood et al., 2013; Coulman et al., 2009; Henry et al., 1998), e.g., Chabrai et al. (2004) have successfully used micro-fabricated silicon MNs for non-viral gene delivery without causing pain. Recently, Zhang et al. (2013) have proposed the use MNs with a gene gun system to assist the micro-particle delivery in the skin. In particular, solid MNs can create holes in the tissue to provide an environment for the penetration of high-speed micro-particles. Furthermore, hollow MNs may allow a number of micro-particles to go through the hollow needles and penetrate to a greater depth into deeper tissue layers to be gene transfected. However, the waste of micro-particles is likely to be higher as the hollow MNs have blockage problems.

There have been a number of studies which report on the effects of MNs types and configuration on drug delivery (Han and Das, 2013; Olatunji et al., 2012; Al-Qallaf and Das, 2009a,b, 2008; Al-Qallaf et al., 2009a,b, 2007; Davidson et al., 2008). The current contribution will focus on the most relevant MNs types and configurations; reviews of other aspects related to the application of MN may be found elsewhere (Olatunji and Das, 2011, 2010).

3.2 MN insertion in skin

MNs can overcome a target surface to provide an advantageous condition for micro-particle delivery (Zhang et al., 2013). Human skin consists of two distinct macroscopic layers called the dermis and the epidermis (Marks et al., 2006; Parker, 1991; Phipps, 1988) which are shown in

Figure 18. The epidermis layer consists of the *stratum corneum* (SC), the *stratum basale* (or *stratum germinativum*), *stratum spinosum*, and the *stratum granulosum* (Gerard et al., 2011; Marks et al., 2006; Holbrook, 1994). The top layer of the skin is the SC which is the major barrier for entry of foreign substances. The thickness of the epidermis varies with gender, age, ethnicity and the regions of the body, but has an average thickness of between 20 and 100 μm (Matteucci et al., 2009; Schaefer and Redelmeier, 1996). The thickness of the *stratum corneum* also varies with the above conditions, but the average thickness is between 10 and 20 μm (Mohammed et al., 2012; Holbrook et al., 1974).

Figure 18: The structure of the skin (MacNeil, 2007)

Normally, MN insertion is painless as it simply penetrates the skin surface without reaching the dermis layer, which contains the nerves ending. (Gupta et al., 2011a; Silpi et al., 2011; Donnelly et al., 2010a; Henry et al., 1998). Pain sensation depends on the MN design which is generally small enough to avoid significant damage to the nerves in the tissue (Shah et al., 2011; Palastanga et al., 2006; Kaushik et al., 2001). Also, as the pain sensation is related to the MN design, an increased size increases the likelihood of stimulating the nerves (Sachdeva and Banga, 2011; Shah et al., 2011; Gupta et al., 2011a,b). There are a number of clinical studies which show that MNs insertions into skin are painless, or the pain is undetectable. For example, Mikolajewska et al. (2010) have used polymeric MN cones with needle height of 644

481 μm , base diameter of 217 μm tapering to tip diameter of 41 μm to progress the skin
 482 pre-treatment and report that the MNs insertion is a painless process. Previously, Haq et al.
 483 (2009) have used several MN arrays with lengths of 180 and 280 μm and compared the pain
 484 responses of a number of subjects for these MNs with the pain responses from hypodermic
 485 needles. They indicated that a hypodermic needle is painful after insertion into skin, and the
 486 pain response of MN insertion is less and decreases with a decrease in needle height.

487
 488 The irregular surfaces and viscoelasticity of skin causes difficulty with MN insertion. In
 489 addition, the skin is generally folded after the insertion of MNs, which may cause MNs to pierce
 490 partially, depending on the MN length (Verbaan et al., 2008). Thus, there is a need to
 491 understand the required force to insert a given MNs in the skin; the depth of MN penetration
 492 into the skin is directly related to the used force for penetration (Olatunji et al., 2013; Donnelly
 493 et al., 2010b). In addition, the depth of MN penetration in the skin is dependent on the length of
 494 MNs (Badran et al., 2009).

495
 496 Figure 19: Methylene blue staining of MN holes on the human skin (needle height: 180 and
 497 280 μm) (Haq et al., 2009)

498
 499 Several studies of MN insertion in skin show that the holes remain in the skin after the removal
 500 of the MNs (e.g., Haq et al., 2009), as shown in Figure 19. Previously, McAllister et al. (2003)
 501 have applied a cylindrical MN of 20 μm diameter to perform staining experiments, which
 502 indicates that a hole will remain after the removal of MNs. Martanto et al. (2004) have shown
 503 that a number of visible holes remained when MNs were applied with needle lengths of 1000

μm and width of $200\ \mu\text{m}$ by $50\ \mu\text{m}$ on a rat skin. In addition, McAllister et al. (2003) have reported a residual MN hole of radius $6\ \mu\text{m}$ following insertion of MNs with radius of $10\ \mu\text{m}$, which means that the holes shrink to ~ 60 percent of the diameter of the MNs. While inserting the MNs, it can also be seen that the entire MN length cannot be inserted into the skin completely, and the depth of penetration of MN in skin is related to its length and the application force. These factors directly affect the size of pierced holes before and after the removal of MNs, Donnelly et al. (2010b) have used optical coherence tomography (OCT) to detect the effect of MN height and application force on the depth of penetration inside the porcine skin. As shown in Figure 20a, the penetration depth is increased significantly by an increase in MN height and application force. In addition, the application force presents a positive effect on the penetration depth of the MN inside skin, as shown in Figure 20b: the pore width is increased by an increase in needle height and application force, when the base width is kept constant. However, there is a clear gap left between the MN base plate and the skin surface. Donnelly et al. (2011) have further used OCT to obtain 3D views of MN embedded in the human skin. They reported that a MN with needle height of $600\ \mu\text{m}$ and base diameter of $300\ \mu\text{m}$ penetrated approximately $460\ \mu\text{m}$ into the human skin with a clear gap of $136\ \mu\text{m}$ between the MN base plate and skin. They also indicated that the width of the pierced holes in the skin was about $265\ \mu\text{m}$ in diameter. These reports demonstrate that MN assisted micro-particle delivery is expected to realize a greater penetration depth of micro-particles in the skin.

4. Potential of MN assisted micro-particle delivery

4.1 MN assisted micro-particle delivery

As mentioned earlier, cell and tissue damages are particular problems for the biolistic gene transfection (Obrien et al., 2011; Uchida et al., 2009; Thomas et al., 2001; Sato et al., 2000; Yoshida et al., 1997). Reduction in the operation pressure and in the particle size can minimize the cell damage, but these also decrease the momentum of the micro-particles. This minimizes the penetration depth of the micro-particles in the tissue and may cause a failure of the DNA transfection.

Figure 20: (a): A 2D OCT image for the investigation of the effect of the MN height on the penetration depth inside the porcine skin (A: 280 μm ; B: 350 μm ; C: 600 μm ; D: 900 μm) (Donnelly et al., 2010b) (b): A 2D OCT image to analyse the effect of the application force on the penetration depth of MN inside the porcine skin (A: 4.4 N; B: 7.0 N; C: 11 N; D: 16.4 N) (Donnelly et al., 2010b) (c): A 3D OCT image showing MN insertion in the skin (needle height: 600 μm ; base width: 300 μm ; spacing: 300 μm) (Donnelly et al., 2011).

Based on a consideration of reducing the operating pressure, a new concept of combining MNs with gene guns for micro-particle delivery has been presented by Zhang et al (2013). As presented in Figure 21, the concept may be developed from the light and Helios gas gun systems. The operating process consists of particle acceleration, separation and deceleration stages. For the acceleration stage, a pellet of micro-particles is loaded onto the ground slide which is accelerated by pressurized gas. The pellet is then separated into individual micro-particles by impaction onto a mesh stopping screen in the separation stage. The separated micro-particles spray forward through a conical nozzle with a uniform velocity and distribution to breach the skin tissue to target the cells of interest.

As shown by Zhang et al (2013), by using an array of MNs it is possible to overcome the effect of the skin on the particle penetration. One of the many advantages of this new concept is that the pressurized gas will be released from the vent holes (see Figure 21), which is likely to avoid the damage of the target from the impaction of pressurized gas on tissue. In addition, the use of the ground slide slows down the velocities of micro-particles to reduce the impact force on tissue to minimize the cell damage. As required, the velocity is controllable by changing the driving pressure of compressed gas, if higher micro-particle velocities are necessary. Even if the micro-particles cannot reach the desired penetration depth due to insufficient momentum,

the use of MNs allows a number of micro-particles to penetrate further to achieve the purpose of gene transfection in the desired depth of the tissue. However, the disadvantage of this concept is that a very high impact velocity of micro-particles is not easy to achieve, because the ground slide has significant mass.

Figure 21: A possible structure of MNs assisted micro-particle delivery system

The solid MN used by Zhang et al. (2013) aimed to create holes on the skin to allow a number of micro-particles penetrate through the pierced holes and increase the penetration depth. Zhang et al. (2013) also indicate that some agglomerated micro-particle may be present in the target; the size of the agglomerates is controllable and decreases with a decrease in mesh size and binder concentration. Results (see Figure 22) have shown that a number of micro-particles are able to penetrate through the pierced holes (created by the MNs) and reached a greater penetration depth inside the target, demonstrating the feasibility of MN assisted micro-particle delivery. In principle, the maximum penetration depth of micro-particles is affected by the particle size, density and operating pressure which determine the momentum of the micro-particles and hence the impact of the particles on the target. An increased length of the pierced holes enhances the particle penetration depth due to a decreased resistance when micro-particle travel in the hole. The length of the pierced holes depends on the height of MN.

581

582

583 Figure 22: Optical microscope image of stainless steel micro-particle penetration into agarose
584 gel (Zhang et al., 2013)

585

586 4.2 Effects of physical approaches to drug delivery

587 With the development of transdermal drug delivery, several physical technologies have been
588 developed, particularly in needle free gene gun systems. In order to understand the
589 advantages of MN assisted micro-particle delivery for the drug/gene delivery, a comparison
590 with other physical cell targeting approaches is presented in this section. Figure 23 illustrates a
591 schematic of four physical cell targeting approaches, which include diffusional delivery (Figure
592 23a), solid MN assisted micro-particle delivery (Figure 23b), hollow MN assisted micro-particle
593 delivery (Figure 23c) and needle free biolistic micro-particle injection (Figure 23d). The route of
594 the diffusion delivery (Figure 23a) is that the molecules permeates through the aperture of the
595 SC and diffuse into the target (Glenn et al., 2003). It is a method which operates without
596 damage to the skin.

597

598 Figure 23: A schematic cross-section of the skin: (a) the normally diffusion route (b) route of
599 solid MN assisted micro-particle delivery (c) route of hollow MN assisted micro-particle delivery
600 (d) route of normal micro-particle penetration using a gene gun system (Kendall, 2006)

601

602 In recent years, needle-free biolistic micro-particle delivery (Figure 23d) provides a great
603 improvement for transdermal gene delivery. The principle of this technique is that DNA is

loaded onto micro-particles which are accelerated to a sufficient velocity to pierce the skin and travel to a certain depth to achieve the DNA transfection in the viable epidermis layer. It can be seen from Figure 23d that micro-particles penetrate to greater depth than diffusion delivery. In addition, biolistic micro-particle delivery is painless as the micro-particles settle within epidermis without reaching the nerves (Quinlan et al., 2001). There are some disadvantages for biolistic micro-particle delivery such as micro-particle penetration causes a significant cell death in the skin due to impaction (O'Brien and Lummis, 2011; Raju et al., 2006; Sato et al., 2000). In addition, pressurized gas may damage the skin surface, if the pressure is over than 200 psi (Belyantseva, 2009; Uchida et al., 2009; Yoshida et al., 1997).

As mentioned earlier, the method of MN assisted micro-particle delivery requires use of a MN to overcome the skin surface to deliver micro-particles to a greater depth. The projected routes are presented in Figure 23b-c. In Figure 23b the penetration depths of micro-particles are greater than for needle-free biolistic micro-particle delivery, due to the pierced holes providing a low resistance path for micro-particle penetration. Figure 23c shows hollow MN assisted micro-particle delivery, which should allow particles to penetrate further in the skin via the hollow needles. However, the disadvantage of MN assisted micro-particle delivery is that the process may be painful if the micro-particles are deliver into dermis which have nerves ending in that layer.

Overall, it is obvious that needle-free biolistic micro-particle injection present more efficiency than diffusion delivery. Ziegler (2008) has shown that acceptable DNA vaccination requires the coated micro-particle to penetrate the skin surface with around 20-100 μm penetration depth. It indicated that needle-free biolistic micro-particle injection achieves a more efficient pharmaceutical effect than diffusion delivery. Further, MN assisted micro-particle delivery may deliver micro-particles deeper than the needle-free gene gun system in the skin to allow deeper tissues to be transfected. This it has been demonstrated that MN assisted micro-particle delivery may achieve a further enhanced DNA transfection in the target.

4.3 Modelling micro-particle delivery in skin

The process of micro-particle delivery is normally divided into two stages, which are the particle acceleration and penetration stages. In the modelling, the acceleration stage should be considered along with the driving source (gas pressure) which accelerates the micro-particle to a sufficient velocity to pierce the skin surface. The velocity varies with the gene gun system design, particle density and size. For the penetration stage, the effect of the skin is the major resistance to prevent the micro-particle delivery. Micro-particle delivery requires breaching of the SC and piercing into the epidermis layer (Yager et al., 2013; Soliman et al., 2011b; Liu, 2006; Quinlan et al., 2001; Bennett et al., 1999; Trainer and Alexander, 1997). The impact velocity, particle size and density, target density and yield stress are the major variables affecting the penetration depth.

Normally, the micro-particle acceleration stage involves gas and particle flow for gas gene gun systems, e.g., CST, PowderJect and Helios Gene gun. The flow is defined as symmetric and fully turbulent in the device (Liu, 2006; Soliman et al., 2011a). A model which has been widely adopted to model micro-particle flow is based on a balance of forces using Newton's second law and Stokes' law. For example, Liu (2006) has focused on simulating the velocity distribution in the converging (conical) section of a venturi system which is developed from a gene gun, namely, the PowderJect system (PowderJect Research Ltd., Oxford, UK) (Bellhouse et al., 1999, 2003, 2006). The particle velocity has been simulated based on a balance between the inertia of micro-particles and other resistance forces acting on the particles. Zhang et al. (2007) have used the MATrix LABoratory (MATLAB, The MathWorks Inc., Natick, USA) (Shampine et al., 1997) to simulate three different stages of the particle delivery in the gene gun, namely, acceleration, separation and deceleration stages. In their work, the particle velocity is analyzed on the basis of Newton's second law in the acceleration stage; an energy conservation law is applied to describe the separation of micro-carriers into micro-particles in the separation stage, and Stokes' law is applied to model the penetration of micro-particles in the deceleration stage into a viscous target. Soliman et al. (2011b) have used a commercial turbo-machinery flow simulator, namely, FINETM/Turbo (NUMECA International, Brussel, Belgium) to simulate the behaviour of gas and particle flow in a supersonic core jet in a gene gun. This work used Newton's second law to mimic the particle

trajectories and determine the penetration depths of micro-particles in the skin. As discussed below, a number of studies have shown that the penetration depth depends on the momentum of micro-particles which again depend on the particle size, density and velocity.

For the penetration stage, various studies have separated the resistance force on the micro-particle into a yield force (F_y), frictional resistive force (F_f) and resistive inertial force of the target material (F_i) (Soliman et al., 2011b; Liu, 2007; Mitchell et al., 2003; Kendall et al., 2001; Dehn, 1987). The force balance equation is shown below:

$$m \frac{dv}{dt} = -(F_i + F_f + F_y) \quad (1)$$

From this the stopping distance can be calculated as

$$d = \frac{4\rho_p r_p}{3\rho_t} \left\{ \ln\left(\frac{1}{2}\rho_t v_i^2 + 3\sigma_y\right) - \ln(3\sigma_y) \right\} \quad (2)$$

Based on this force balance equation, the theoretical penetration depth (equation 2) is obtained and adopted for the modelling. This model has been widely adopted in several studies. For example, Soliman et al. (2011b) have modelled the delivery of golden particle of diameters 1.8 and 5 μm using 3 MPa pressure and have shown that penetration depths of 95 μm and 135 μm can be achieved for particle of diameters 1.8 and 5 μm , respectively. Kendall et al. (2001) have analysed the golden particle penetration by using equation 2 for particle penetration in human and porcine skins. In addition, predictions from the theoretical model have been shown to agree well with the experimental results by Kendall et al. (2001). More recently, Soliman et al. (2011b) have also implemented equation 2 in a theoretical model which is implemented using FINE/Turbo code to calculate the penetration depth of gold micro-particle inside the skin.

5. Conclusion

The background of the gene gun system for micro-particle delivery is reviewed in this paper. A number of gene gun systems have been listed and the operating principles along with their advantages and disadvantages have been studied briefly. In addition, the recommended gas type, particle material and size for these type engineering systems are discussed. The range of particle velocities and applied operating pressures for several gene gun systems are described, which indicated that cell/tissue damage is a major problem for biolistic

micro-particle delivery, due to the impaction of pressurized gas and high-speed micro-particles on the target tissue. In addressing this point, a new concept is proposed of MN assisted micro-particle delivery, which combines a gene gun system with MN to enhance the penetration depth of micro-particles. This technique may reduce the cell damage from pressurized gas and reduce the impact velocity of micro-particle. In order to further understand the MN assisted micro-particle delivery, a number of MN designs have been discussed, paying attention to key characteristics that affect biolistic delivery.

A number of researchers have studied the MN insertion in the skin which suggests that holes remained on the skin after the removal of MNs. These indirectly show the feasibility of the MN assisted micro-particle delivery to enhance the penetration depths of micro-particles inside the target. Based on the above research works, the detailed penetration route of MN assisted micro-particle delivery is also discussed and compared with other physical approaches on drug delivery. Finally, various models of micro-particle delivery for different gene gun systems are described to understand the theoretical principles of micro-particle penetration and which may be used for modelling of MN assisted micro-particle delivery.

6. Conflict of Interest

Authors declare no conflict of interest

7. References

Ahlen G., Sallberg M., Frelin L., (2013). Methods for monitoring gene gun-induced HBV- and HCV-specific immune responses in mouse models, *Methods Mol. Biol.*, **940**: 239-267

Al-Qallaf B. and Das D.B., (2009a). Optimizing microneedle arrays for transdermal drug delivery: Extension to non-square distribution of microneedles, *Journal of Drug Targeting*, **17**(2), pp.108-122, DOI: 10.1080/10611860802472370.

Al-Qallaf B. and Das D.B., (2009b). Optimizing Microneedle Arrays to increase Skin Permeability for Transdermal Drug Delivery, *Annals of the New York Academy of Sciences*, **1161**: 83-94, DOI: 10.1111/j.1749-6632.2009.04083.x.

726 Al-Qallaf B., Das D.B., Davidson A., (2009a). Transdermal Drug Delivery by Coated
 727 Microneedles: Geometry Effects on Drug Concentration in Blood, *Asia-Pacific Journal of*
 728 *Chemical Engineering*, **4**(6): 845-857, DOI: 10.1002/apj.353.

729

730 Al-Qallaf B., Mori D., Olatunji O., Das D.B., Cui Z.F., (2009b). Transdermal Drug Delivery by
 731 Microneedles: Does Skin Metabolism Matter?, *International Journal of Chemical Reactor*
 732 *Engineering*, **7**: A69.1-A69.23.

733

734 Al-Qallaf B. and Das D.B., (2008). Optimization of square microneedle arrays for increasing
 735 drug permeability in skin, *Chemical Engineering Science*, **63**: 2523-2535, DOI:
 736 10.1016/j.ces.2008.02.007.

737

738 Al-Qallaf B., Das D.B., Mori D., Cui Z.F., (2007). Modelling transdermal delivery of high
 739 molecular weight drugs from microneedle systems, *Philosophical Transactions of The Royal*
 740 *Society A*, **365**: 2951-2967, DOI: 10.1098/rsta.2007.0003.

741

742 Aravindaram K., Yang N.S., (2009). Gene gun delivery systems for cancer vaccine
 743 approaches, *Methods Mol. Biol.*, **542**: 167-178

744

745 Armaleo D., Ye G.N., Klein T.M., Shark K.B., Sanford J.C., Johnston S.A., (1992). Biolistic
 746 nuclear transformation of *Saccharomyces cerevisiae* and other fungi, *Curr. Genet.*, **17**:97-103

747

748 Arora A., Prausnitz M.R., Mitragotri S., (2008) Micro-scale devices for transdermal drug
 749 delivery, *Int. J. Pharm.* **364**:227-236

750

751 Badran, M. M., Kuntsche, J., Fahr, A., (2009). Skin Penetration Enhancement by a
 752 Microneedle Device (Dermaroller®) *in vitro*: Dependency on Needle Size and Applied
 753 Formulation, *Pharm Sci.*, **36**: 511- 523.

754

755 Bal S.M., Caussin J., Pavel S., Bouwstra J.A. (2008). *In vivo* assessment of safety of
 756 microneedle arrays in human skin, *European Journal of Pharmaceutical Sciences*, **35**(3):
 757 193-202.

758

759 Bastian S., Busch W., Kuhnel D., Springer A., Holke R., Scholz S., Pompe W., Gelinsky M.,
760 Potthoff A., Richter V., Ikonomidou C., Schirmer K., (2009). Toxicity of tungsten carbide and
761 cobalt-doped tungsten carbide nanoparticles in mammalian cells *in vitro*, *Environ. Health*
762 *Perspect.*, **117**(4): 530-536

763

764 Bellhouse B.J., Sarpie D.F., and Greenford J.C., (1994). Needleless syringe using
765 supersonic gas flow for particle delivery, Patent: US5899880 A, *United States Patent*

766

767 Bellhouse B.J., Bell J., Millward H.R., Phillips M.J., Nabulsi S.M., (2003). Particle delivery, US
768 Patent: US 6592545 B1, *United States Patent*

769

770 Bellhouse B.J., Greenford J.C., Sarpie D.F., (2006). Particle delivery, Patent: EP 1637173 A2
771

772 Belyantseva I.A., (2009). Helios Gene Gun–Mediated Transfection of the Inner Ear Sensory
773 Epithelium, *Auditory and Vestibular Research: Methods and Protocols*, Chapter 7, Editor
774 Sokolowski B., **493**: 103-123

775

776 Bennett A.M., Phillpotts R.J., Perkins S.D., Jacobs S.C., Williamson E.D., (1999). “Gene gun
777 mediated vaccination is superior to manual delivery for immunisation with DNA vaccines
778 expressing protective antigens from *Yersinia pestis* or Venezuelan Equine Encephalitis virus”,
779 *Vaccine*, **18**(7–8): 588–596.

780

781 Bio-Rad Laboratories, (2013). USA. (also see, <http://www.bio-rad.com/>)

782

783 Binyamin G., Shafi B.M., Mery C.M. (2006). Biomaterials: A primer for surgeons, *Seminars in*
784 *Pediatric Surgery*, **15**(4): 276-283.

785

786 Biopex, (2013). SJ-500 Portable gene gun
787 (<http://www.biopex.com/SJ-500-Portable-gene-gun-28-10.html>)

788

- Bryan M., Guyach S., Norris K.A., (2013). Biolistic DNA vaccination against trypanosome infection, *Methods Mol. Biol.*, **940**: 305-315
- Cao Y., Zhao B., Han Y.H., Zhang J., Li X.Z., Qiu C.H., Wu X.J., Hong Y., Ai D.Z., Lin J.J., Fu Z.Q., (2013). Gene gun bombardment with DNA-Coated Golden particles enhanced the protective effect of a DNA vaccine based on thioredoxin glutathione reductase of schistosoma japonicum, *BioMed. Research International*, **2013**: 952416, DOI: 10.1155/2013/952416.
- Chabrai F., Bouris K., Jones T., Barrow D., Allender C., Hann A., Brain K., Birchall J., (2004). "Microfabricated silicon microneedles for nonviral cutaneous gene delivery". *British Journal of Dermatology*, **150**(5): 869-877
- Chen B., Wei J., Tay F.E.H., Wong Y.T., Iliescu C., (2008). Silicon microneedle array with biodegradable tips for transdermal drug delivery, *Microsystem Technologies*, **14**: 1015-1019.
- Chen D., Maa Y., Haynes J.R., (2002). Needle-free epidermal powder immunization. *Expert Rev. Vaccines*, **1**(3): 89-100
- Chen P.C., Hsieh S.J., Chen C.C., Zou J., (2013). A three-dimensional enormous surface area aluminium microneedle array with nanoporous structure, *Journal of Nanomaterials*, doi: [10.1155/2013/164953](https://doi.org/10.1155/2013/164953) (in press)
- Christou P., McCabe D.E., Swain W.F., (1990). "Soybean genetic-engineering commercial production of transgenic plants", *Trends in biotechnology*, **8**: 145 -151
- Coulman S., Allender C., Birchall J., (2006). "Microneedles and other physical methods for overcoming the stratum corneum barrier for cutaneous gene therapy", *Ther. Drug Carrier Syst.*, **23**(3): 205-258

818 Da'dara A.A., Skelly P.J., Fatakdawala M., Visovatti S., Eriksson E., Harn D.A., (2002).
819 Comparative efficacy of the schistosoma mansoni nucleic acid vaccine, sm23, following
820 microseeding or gene gun delivery, *Parasite Immunology*, **24**: 179-187
821

822 Davis, S.P., Prausnitz, M.R., Allen, M.G., (2003). Fabrication and characterization of Laser
823 Micromachined Hollow Microneedles. *The 12th Int. Conf. Solid states Sensor, Actuators and*
824 *Microsystems*. Boston.
825

826 Davidson A., Al Qallaf B., Das D.B., (2008). Transdermal Drug Delivery by Coated
827 Microneedles: Geometry Effects on Effective Skin Thickness and Drug Permeability, *Chemical*
828 *Engineering Research and Design*, **86**: 1196-1206, DOI: 10.1016/j.cherd.2008.06.002.
829

830 Davis S.P., Landis B.J., Adams Z.H., Allen, M.G., Prausnitz, M.R., (2004). Insertion of
831 microneedles into tissue: measurement and prediction of insertion force and needle fracture
832 force, *Journal of Biomechanics*, **37**: 1155-1163
833

834 Dehn J., (1987). A Unified theory if penetration, *International journal of impact engineering*, **5**:
835 239-248
836

837 Disegi J.A., Eschbach L., (2000). Stainless steel in bone surgery, *Injury*, **31**(4): 2-6
838

839 Donnelly R.F., Raj S.T.R., Woolfson A.D., (2010a). Microneedle-based drug delivery
840 systems: microfabrication, drug delivery, and safety, *Drug Deliv.*, **17**(4): 187-207
841

842 Donnelly R.F., Garland M.J., Morrow D.I.J., Migalska K., Singh T.R.R., Majithiya R., Woolfson
843 D.A., (2010b). Optical coherence tomography is a valuable tool in the study of the effects of
844 microneedle geometry on skin penetration characteristics and in-skin dissolution, *Journal of*
845 *Controlled Release*, **147**: 333-341.
846

847 Donnelly R.F., Majithiya R., Singh T.R.R., Morrow D.I.J., Garland M.J., Demir Y.K., Migalska
848 K., Ryan E., Gillen D., Scott C.J., Woolfson A.D., (2011). Design, optimization and

characterisation of polymeric microneedle arrays prepared by a novel laser-based micromoulding technique, *Pharm. Res.*, **28**: 41-57

Donnelly R.F., Garland M.J., Singh T.R.R., Migalska K., Majithiya R., McCrudden C.M., Kole P.L., Mahmood T.M.T., McCarthy H.O., Willfson A.D., (2012). Hydrogel-Forming microneedle array for enhanced transdermal drug delivery, *Adv. Funct. Mater.*, **22**:4879-4890

Ettinger M., Peckl S.D., Gruber C., Laimer M., Thalhamer J., Hintner H., Gratz I.K., Bauer J.W., (2012). Transcutaneous gene gun delivery of hNC16A induces BPAG2-Specific tolerance, *Journal of Investigative Dermatology*, 132(6):1665-1671

Fernandez L.J., Altuna A., Tijero M., Gabriel G., Villa R., Rodriguez M.J., Batlle M., Vilares R., Berganzo J., Blanco F.J., (2009). Study of functional viability of SU-8-based microneedles for neural applications, *J. Micromech. Microeng.* **19**: doi:10.1088/0960- 1317/19/2/025007.

Gattiker G.E., Kaler K.V.I.S., Mintchev M.P., (2005). "Electronic mosquito: designing a semi-invasive microsystem for blood sampling, analysis and drug delivery applications", *Microsystem Technologies*, **12**(1-2): 44–51

Gardeniers H.J.G.E., Luttge R., Berenschot E.J.W. de Boer M.J., Yeshurun S.Y., Hefetz M., van't Oever R., van den Berg A., (2003). "Silicon Micromachined Hollow Microneedles for Transdermal Liquid Transport". *Journal of Microelectromechanical Systems*, **12**(6): 855-862

Glenn G.M., Kenney R.T., Ellingsworth L.R., Frech S.A., Hammond S.A., Zoetewewei J.P., Transcutaneous immunization and immunostimulant strategies: capitalizing on the immunocompetence of the skin, *Expert Rev., Vaccines*, **2**(2): 253-267

Gill, H.S., Denson, D.D., Burris, B.A., Prausnitz, M.R., (2008). Effect of Microneedle design on pain in Human subjects, *Clin. J. Pain.*, **24**:585-594.

879 Gu B., Kale S.D., Wang Q.H., Wang D.H., Pan Q.N., Cao H., Meng Y.L., Kang Z.S., Tyler B.M.,
 880 Shan W.X., (2011). Rust secreted protein Ps87 is conserved in diverse fungal pathogens and
 881 contains a RXLR-like motif sufficient for translocation into plant cells, *PLoS One.*, **6**(11):
 882 e27217
 883
 884 Gupta J., Park S.S., Bondy B., Felner E.I., Prausnitz M.R., (2011a). Infusion pressure and pain
 885 during microneedle injection into skin of human subjects, *Biomaterials*, **32**: 6823-6831
 886
 887 Gupta J., Gill H.S., Andrews S.N., Prausnitz M.R., (2011b). Kinetics of skin resealing after
 888 insertion of microneedles in human subjects, *Journal of Controlled Release*, **154**: 148-155
 889
 890 Han M., Hyun D.H., Park H.H., Lee S.S.L., Kim C.H., Kim C.G., (2007). A novel fabrication
 891 process for out-of-plane microneedle sheets of biocompatible polymer, *J. Micromech.*
 892 *Microeng.*, **17**: 1184-1191.
 893
 894 Han R., Peng X., Reed C.A., Cladel N.M., Budgeon L.R., Pickel M.D., Christensen N.D.,
 895 (2002). Gene gun-mediated intracutaneous vaccination with papillomavirus E7 gene delays
 896 cancer development of papillomavirus-induced skin papillomas on rabbits, *Cancer Detect*
 897 *Prev.*, **26**(6): 458-467
 898
 899 Han T., Das D.B., (2013). Permeability enhancement for transdermal delivery of large
 900 molecule using low-frequency sonophoresis combined with microneedles, *Journal of*
 901 *Pharmaceutical Science*, **102**(10): 3614-3622
 902
 903 Han X., Wang D.K., Wang Y., (2008). "Development of Transdermal Delivery System of
 904 Microneedles", *Chinese Journal of Pharmaceutics*, **6**(5): 296-300
 905
 906 Haq M.I., Smith E., John D.N., Kalavala M., Edwards C., Anstey A., Morrissey A., Birchall J.C.,
 907 (2009). Clinical administration of microneedles: skin puncture, pain and sensation, *Biomed.*
 908 *Microdevices*, **11**: 35-47
 909

910 Hardy M.P., Kendall M.A.F., (2005). Mucosal deformation from an impinging transonic gas jet
 911 and the ballistic impact of microparticles, *Phys. Med. Biol.*, **50**(19): 4567-4580
 912

913 Henry S., McAllister D.V., Allen M.G., Prausnitz M.R., (1998). Micromachined needles for the
 914 transdermal delivery of drugs, *Proc. IEEE Micro Electro Mechanical Systems*, MEMS'98,
 915 Heidelberg Germany, **25**(29): 494–498
 916

917 Henry, S., McAllister, D.V., Allen, M.G., Prausnitz, M.R. (1998). Microfabricated microneedles:
 918 a novel approach to transdermal drug delivery, *J. Pharm. Sci.*, **87**: 922-925.
 919

920 Holbrook, K.A. (1994). Ultrastructure of the Epidermis: Chapter 1 of the Keratinocyte
 921 Handbook, edited by Leigh, I.M. and Lane, E.B., Watt, F.M., Cambridge, Great Britain,
 922 *Cambridge University Press*. : 3–39.
 923

924 Holbrook, K. A., Odland, G.F. (1974). Regional Differences in the thickness (cell layers) of the
 925 human stratum corneum: an ultrasound analysis, *J Invest Dermatol.*, **62**: 415-422.
 926

927 Hou L.Y., Wang P.Y., Kong F.T., Park H., Kobihiro K. and Ohama T., (2013). Mesoporous TiO₂
 928 nanoparticles: A new material for biolistic bombardment, *Phycological Research*, **61**: 58-60
 929

930 Huang P.H., Chen P.Y., (2011). Design of a two-stage electromagnetic impulse force circuit for
 931 gene gun, *Journal of Marine Science and Technology*, **19**(6): 686-692
 932

933 Ikemoto K., Sakata I., Sakai T., (2012). Collision of millimetre droplets induces DNA and
 934 protein transfection into cells, *Scientific Reports*, **2**: 289
 935

936 Jiang, J., Moore, J.S., Edelhauser, H.F., Prausnitz, M. R. (2009). Intrasccleral Drug Delivery to
 937 the Eye Using Hollow microneedles, *Pharm Re.s*, **26**: 395–403.
 938

939 Jiang J, Gill HS, Ghate D, McCarey BE, Patel SR, Edelhauser HF, Prausnitz MR. (2007).
 940 Coated microneedles for drug delivery to the eye, *Invest. Ophthalmol. Vis. Sci.*, **48**: 4038–43.

941

942 Kaushik S., Hord A.H., Denson D.D., McAllister D.V., Smitra S., Allen M.G., Prausnitz M.R.,
 943 (2001). Lack of pain associated with microfabricated microneedles, *Anesth. Analg.*, **92**(2):
 944 502-504

945

946 Kendall M.A.F., Carter F.V., Mitchell T.J., and Bellhouse B.J., (2001). Comparison of the
 947 transdermal ballistic delivery of micro-particle into human and porcine skin, *Proceedings of the*
 948 *23rd Annual International Conference of the IEEE*: 2991-94

949

950 Kendall M.A.F., (2002). The delivery of particulate vaccines and drugs to human skin with a
 951 practical, hand –held shock tube-based system, *Shock Waves*, **12**: 23-30

952

953 Kendall M.A.F., Quinlan N.J., Thorpe S.J., Bellhous B.J., Ainsworth R.W. (2004a).
 954 Measurements of the gas and particle flow within a converging-diverging nozzle for high speed
 955 powdered vaccine and drug delivery, *Exp. Fluids*, **37**: 128-36

956

957 Kendall M.A.F., Rishworth S., Carter F., Mitchell T., (2004b). Effects of relative humidity and
 958 ambient temperature on the ballistic delivery of micro-particles to excised porcine skin, *J.*
 959 *Invest. Dermatol.* **122**:739-746

960

961 Kendall M.A.F., Mitchell T.J., Costigan G., Armitage M., Lenzo J.C., Thomas J.A., Garnier
 962 C.V., Zosky G.R., Turner D.J., Stumables P.A., Sly P.D., Holt P.G., Thomas W.R., (2006).
 963 Downregulation of IgE antibody and allergic responses in the lung by epidermal biolistic
 964 microparticle delivery, *J. Allergy Clin. Immunol.*, **117**: 275-282

965

966 Kendal M.A.F., (2006). Engineering of needle-free physical methods to target epidermal cells
 967 for DNA vaccination, *Vaccine*, **24**: 4651-4656

968

969 Kim Y.C., Quan F.S., Compans R.W., Kang S.M., Prausnitz M.R., (2010). Formulation of
 970 microneedles coated with influenza virus-like particle vaccine, *AAPS PharmSciTech.*, **11**(3):
 971 1193-1201

972

973 Kim Y.C., Park J.H., Prausnitz M.R., (2012). Microneedles for drug and vaccine delivery,
 974 *Advanced Drug Delivery Reviews*, **64**: 1547-1568

975

976 Klein T.M., Wolf E.D., Wu R., Sanford J.C., (1987). High-velocity microprojectiles for delivery
 977 of nucleic acids into living cells, *Nature(London)*, **327**: 70-73

978

979 Klein T.M., Arentzen R., Lewis P.A., McElligott F.S., (1992). Transformation of microbes,
 980 plants by particle bombardment, *Biotechnology*, 10: 286-291

981

982 Koelmans W.W., Krishnamoorthy G., Heskamp A., Wissink J., Misra S., Tas N., (2013).
 983 Microneedle Characterization using a double-layer skin simulant, *Mechanical Engineering*
 984 *Research*, **3**(2): 51-63

985

986 Kuriakose B., Toit E.S., Jordaan A., (2012). Transient gene expression assays in rose tissues
 987 using a Bio-Rad Helios hand-held gene gun, *South African Journal of Botany*, **78**: 307-311

988

989 Kuriyama S., Mitoro A., Tsujinoue H., Nakatani T., Yoshiji H., Tsujimoto T., Yamazaki M.,
 990 Fukui H., (2000). Particle-mediated gene transfer into murine livers using a newly developed
 991 gene gun, *Gene Ther.*, **7**: 1132-1136.

992

993 Liu Y., Kendall M.A.F., (2004a). Numerical study of a transient gas and particle flow in a
 994 high-speed needle-free ballistic particulate vaccine delivery system, *Journal of Mechanics in*
 995 *Medicine and Biology*, **4**(4): 559-578

996

997 Liu Y., Kendall M.A.F., (2004b). Numerical simulation of heat transfer from a transonic jet
 998 impinging on skin for needle-free powdered drug and vaccine delivery, Proceedings of the
 999 institution of mechanical engineers, part C: *J. Mech, Eng, Sci.*, **218**(11): 1373-1383

1000

1001 Liu Y., Kendall M.A.F., (2006). Numerical analysis of gas and micro-particle interactions in a
 1002 hand-held shock tube device, *Biomed Microdevices*, **8**: 341-351

1003

1004 Liu Y., (2006). Physical-mathematical modeling of fluid and particle transportation for DNA
 1005 vaccination, *International Journal of Engineering Science*, **44**(15-16): 1037-1049
 1006

1007 Liu Y., (2007). Impact studies of high-speed micro-particle following biolistic delivery, *IEEE*
 1008 *TRANSACTIONS ON BIOMEDICAL ENGINEERING*, **54**(8): 1507-1513
 1009

1010 Liu Y., Truong N.K., Kendall M.A.F., Bellhouse B.J., (2007). Characteristics of a micro-biolistic
 1011 system for murine immunological studies, *Biomed. Microdevices*, **9**:465-474
 1012

1013 Liu Y., Costigan G., Bellhouse B.J., (2008). Swirling effects on the performance of the
 1014 micro-particle acceleration and penetration: parametric studies, *Powder Technol.* **183**:
 1015 189-195
 1016

1017 Liu Y., (2008). Simulation of transient shock motion within a biological contoured shock tube
 1018 system, *Physica D: Nonlinear Phenomena*, **237**(2): 233-242
 1019

1020 Macklin, M.D., Drape, R.J., Swain, W.F., (2000). Preparation for particle-mediated gene
 1021 transfer using the Accell gene gun, *Methods in molecular medicine*, **29**: 297-303
 1022

1023 MacNeil S., (2007). Progress and opportunities for tissue-engineered skin, *Nature*, **445**:
 1024 874-880
 1025

1026 Manjila S.B., Baby J.N., Bijin E.N., Constantine I., Pramod K., Valsalakumari J., (2013). Novel
 1027 gene delivery system, *International Journal of Pharmaceutical Investigation*, **3**(1): 1-7
 1028

1029 Marks J.G., Miller J.J., Lookingbill D.P., (2006). Principles of Dermatology, *4 editions*,
 1030 Saunders Elsevier: 1-7, ISBN: 1416031855
 1031

1032 Marrion M.C., Gostigan G., Bellhouse B.J., (2005). Flow separation control in over-expanded
 1033 supersonic nozzle, *35th AIAA fluid dynamics conference and exhibit*.

1034

1035 Martanto W., Davis S.P., Holiday N.R., Wang J., Gill H.S., Prausnitz M.R. (2004). Transdermal

1036 delivery of Insulin using microneedles *in vivo*, *Pharmaceutical Research*, **21**: 947-952.

1037

1038 Matteucci M., Fanetti M., Casella M., Gramatica F., Gavioli L., Tormen M., Grenzi G., De

1039 Angelis F., Di Fabrizio E. (2009). Poly vinyl alcohol re-usable masters for microneedle

1040 replication. *Microelectronic Engineering*, **86**(4-6): 752-756, doi:10.1016/j.mee.2009.01.068

1041

1042 Matthews K., Bailey S.L., Gossner A.G., Watkins C., Dalziel R.G., Hopkin J., (2007). Gene

1043 gun-delivered pGM-CSF adjuvant induces enhanced emigration of two dendritic cell subsets

1044 from the skin, *Scand. J. Immunol.*, **65**(3):221-229

1045

1046 Mcallister D.V., Allen M.G., (2000). Prausnitz MR. Microfabricated microneedles for gene and

1047 drug delivery, *Annual Review Biomedical Engineering*, **2**: 289 –313.

1048

1049 McAllister D.V., Wang P.M., Davis S.P., Park J.H., Canatella P.J., Allen M.G., Prausnitz, M.R.

1050 (2003). Microfabricated needles for transdermal delivery of macromolecules and

1051 nanoparticles: fabrication methods and transport studies. *PNAS*, **100**: 13755–13760.

1052

1053 Meacham J.M., Durvasula K., Degertekin F.L., Fedorov A.G., (2013). Physical methods for

1054 intracellular delivery: Particle aspects from laboratory use to industrial-scale processing,

1055 *Journal of Laboratory Automation*, DOI: 10.1177/2211068213494388

1056

1057 Memon S., Pathan D.N., Ziyaurrahman A.R., Bagwan A., Sayed B., (2011). Microneedle as a

1058 novel drug delivery system: A review, *Internation Research Journal of Pharmacy*, **2**(2): 72-77

1059

1060 Menezes V., Mathew Y., Takayama K., Kanno A., Hosseini H., (2012). Laser plasma jet driven

1061 microparticles for DNA/Drug delivery, *PLoS ONE*, **7**(11): e50823

1062

1063 Mikolajewska P., Donnelly R.F., Garland M.J., Morrow D.I.J., Singh T.R.R., Lani V., Moan J.,

1064 Juzeniene A., (2010). Microneedle pre-treatment of human skin improves 5-aminolevulinic

1065 acid (ALA) and 5-aminolevulinic acid methyl ester (MAL) induced PpIX production for topical
 1066 photodynamic therapy without increase in pain or erythema, *Pharm. Res.*, **27**: 2213-2220
 1067

1068 Mikszta J.A., Alarcon J.B., Brittingham J.M., Sutter D.E., Pettis R.J., Harvey N.G., (2002).
 1069 Improved genetic immunization via micromechanical disruption of skin-barrier function and
 1070 targeted epidermal delivery, *Nat. Med.*, **8**: 415–419.
 1071

1072 Mitchell T.J., Kendall M.A.F., Bellhouse B.J., (2003). A ballistic study of micro-particle
 1073 penetration to the oral mucosa, *International Journal of Impact Engineering*, **28**: 581-599
 1074

1075 Mohammed D., Matts P.J., Hadgraft J., Lane M.E., (2012). Variation of stratum corneum
 1076 biophysical and molecular properties with anatomic site, *The American Association of*
 1077 *Pharmaceutical Scientists*, **14**(4): 806-812
 1078

1079 Nagata T., Koide Y., (2010). Induction of specific CD8⁺ T cells against intracellular bacteria by
 1080 CD 8⁺ T-cell-oriented immunization approaches, *Journal of Biomedicine and Biotechnology*,
 1081 **2010**: 764542, DOI: 10.1155/2010/764542.
 1082

1083 Nayak A., Das D.B., (2013). Potential of biodegradable microneedles as a transdermal
 1084 delivery vehicle for lidocaine, *Biotechnology Letters*, **35**(9): 1351-1363
 1085

1086 Nguyen-Hoai T., Hohn O., Vu M.D., Baldenhofer G., Sayed A.M.S., Dorken B., Norley S., Lipp
 1087 M., Pezzutto A., Westermann J., (2012). CCL19 as an adjuvant for intradermal gene gun
 1088 immunization in a Her2/neu mouse tumor model: improved vaccine efficacy and a role for B
 1089 cells as APC, *Cancer Gene Ther.*, **19**(12): 880-887
 1090

1091 O'Brien J.A., Lummis S.C.R., (2011). Nano-biolistics: a method of biolistic transfection of cells
 1092 and tissues using a gene gun with novel nanometer-sized projectiles, *BMC Biotechnology*, **11**;
 1093 66
 1094

1095 O'Brien J.A., Lummis S.C.R., (2007). Diolistics: incorporating fluorescent dyes into biological
 1096 samples using a gen gun, *Trends in Biotechnology*, **25**(11): 530-534
 1097

1098 O'Brien J.A., Lummis S.C.R., (2006). Biolistic transfection of neuronal cultures using a
 1099 hand-held gene gun, *Nat. Protoc.* **1**(2): 977-981
 1100

1101 Oh, j., Park, H., Do, K., Han, M., Hyun, D., Kim, C., Kim,C., Lee, S., Hwang, S., Shin, S., Cho,
 1102 C. (2008). Influence of the delivery systems using a microneedle array on the permeation of a
 1103 hydrophilic molecule, calcein. *European Journal of Pharmaceutics and Biopharmaceutica*. **69**:
 1104 1040-1045.
 1105

1106 Olatunji, O., Das, D.B., (2010). Painless Drug delivery using microneedles. Current
 1107 technologies to increase the transdermal delivery of drugs (Editor: Joan Escobar Chavez).
 1108 *Bentham Science Publishers* (available online at
 1109 http://www.benthamdirect.org/pages/b_getarticlebybook.php). ISBN:978-1-60805-191-5
 1110

1111 Olatunji, O., Das, D.B., (2011). Drug delivery using Microneedles. Comprehensive
 1112 Biotechnology (Editor: Zhanfeng Cui). 2nd Edition, MS number 501. MRW, Elsevier, The
 1113 Boulevard, Langford Lane, Oxford, United Kingdom. ISBN: 13: 978-0-444-53352-4
 1114

1115 Olatunji, O, Das, DB, Nassehi, V (2012) Modelling transdermal drug delivery using
 1116 microneedles: Effect of geometry on drug transport behaviour, *Journal of Pharmaceutical*
 1117 *Sciences*, **101**(1): 164-175, ISSN: 0022-3549. DOI: 10.1002/jps.22736.
 1118

1119 Olatunji O., Das D.B., Garland M.J., Belaid L., Donnelly R.F., (2013). Influence of Array
 1120 interspacing on the force required for successful microneedle skin penetration: theoretical and
 1121 practical approaches, *Journal of pharmaceutical science*, **102**(4): 1209-1221
 1122

1123 Palastanga N., Field D., Soames R.W., (2006). Anatomy and human Movement: Structure and
 1124 function, 5th edition, London: *Elsevier Health Science*, ISBN-13: 9780750688147
 1125

1126 Park J.H., Allen M.G., Prausnitz M.R., (2005). Biodegradable polymer microneedles:
 1127 fabrication, mechanics and transdermal drug delivery, *Journal of Controlled Release*, **104**(1):
 1128 51-66.
 1129
 1130 Park J.H., Allen M.G., Prausnitz M.R., (2006). Polymer microneedles for controlled-release
 1131 drug delivery, *Pharm. Res.*, **23**: 1008-1019.
 1132
 1133 Parker E.R., Rao M.P., Turner K.L., Meinhart C.D., MacDonald N.C. (2007). Bulk
 1134 micromachined titanium microneedles, *Journal of Microelectromechanical System*, **16**:
 1135 289-295.
 1136
 1137 Parker, F. (1991). Structure and Function of the Skin: Chapter 1 in Dermatology edited by
 1138 Orkin, M. and Maibach, H. I. and Dahl, M. V., *1st edition* (Norwalk, Connecticut: Appleton and
 1139 Lange).
 1140
 1141 Phipps, J. B. Padmanabhan, R. V. Lattin, G. A. (1988). Transport of Ionic Species Through
 1142 Skin. *Solid State Ionics*, **28**: 1778–1783.
 1143
 1144 Prausnitz M.R., Langer R., (2008). Transdermal drug delivery, *Nat. Biotech.*, **26**(11):
 1145 1261-1268
 1146
 1147 Qi P., Chen G.G., Wei P., (2007). “Research and Development of Microneedles for
 1148 Transdermal Drug Delivery”, *Journal of Nanjing University of Technology (Natural Science*
 1149 *Edition)*, **29**(4): 57-63
 1150
 1151 Quan F.S., Kim Y.C., Compans R.W., Prausnitz M.R., Kang S.M., (2010). Does sparing
 1152 enabled by skin immunization with influenza virus-like particle vaccine using microneedles, *J.*
 1153 *Control Release*, **147**(3): 326-332
 1154

1155 Quinlan N.J., Kendall M.A.F., Bellhouse B.J., Ainsworth R.W., (2001). Investigations of gas
 1156 and particle dynamics in first generation needle-free drug delivery device, *Shock Waves*, **10**:
 1157 395-404
 1158
 1159 Raju P.A., Truong N.K., Kendall M.A.F., (2006). Assessment of epidermal cell viability by near
 1160 infra-red two-photon microscopy following ballistic delivery of gold micro-particles, *Vaccine*,
 1161 **24**: 4644-4647
 1162
 1163 Rao W., (2010). A proteomic analysis of lipid raft and GPI anchored protein in *Caenorhabditis*
 1164 *elegans*, University of Leeds (UK), *Thesis for the degree of Doctor of Philosophy*
 1165
 1166 Rasel M.A.I., Taher M.A., Kim H.D., (2013). A study on the gas-solid particle flows in a
 1167 needle-free drug delivery device, *Journal of Thermal Science*, **22**(4): 340-344
 1168
 1169 Rinberg D., (2005). Pneumatic capillary gun for ballistic delivery of micro-particles, *Applied*
 1170 *Physics Letters*, **87**: 014103
 1171
 1172 Rosi N.L., Giljohann D.A., Thaxton C.S., Lytton-Jean A.K., Han M.S., Mirkin C.A., (2006).
 1173 Oligonucleotide-modified gold nanoparticles for intracellular gene regulation, *Science*, **312**:
 1174 1027-1030
 1175
 1176 Russell J.A., Roy M.K., Sanford J.C., (1992). Physical trauma and tungsten toxicity reduce the
 1177 efficiency of biolistic transformation, *Plant Physiol.*, **98**: 1050-1056
 1178
 1179 Sachdeva V., Banga A.K., (2011). Microneedles and their applications, *Recent Pat. Drug*
 1180 *Delivery Formulation*, **5**: 95-132
 1181
 1182 Sakai T., Hisaeda H., Nakano Y., Ishikawa H., Maekawa Y., Ishii K., Nitta Y., Miyazaki J.,
 1183 Himeno K., (2000). Gene gun-mediated delivery of an interleukin-12 expression plasmid
 1184 protects against infections with the intracellular protozoan parasites *leishmania major* and
 1185 *Trypanosoma cruzi* in mice, *Immunology*, **99**: 615-624

1186

1187 Sanford J.C., Klein T.M., Wolf E.D., Allen N., (1987). Delivery of substances into cells and
 1188 tissues using a particle bombardment process, *Particulate Science and Technology*, **5**(1):
 1189 27-37

1190

1191 Sanford J.C., Smith F.D., Russell J.A., (1993). Optimizing the biolistic process for different
 1192 biological applications, *Methods Enzymol.*, **217**: 483-509

1193

1194 Satish D., (2009). *In vitro* manipulation of pollen for transformation in cotton and tomato,
 1195 University of Agricultural Sciences (India), *Thesis for the degree of Doctor of Philosophy*.

1196

1197 Sato H., Hattori S., Kawamoto S., Kudoh I., Hayashi A., Yamamoto I., Yoshinari M., Minami
 1198 M., Kanno H., (2000). *In vivo* gene gun-mediated DNA delivery into rodent brain tissue,
 1199 *Biochem. Biophys. Res. Commun.* **270**(1): 163-170.

1200

1201 Schaefer H., Redelmeier T.E. (1996). Skin barrier: Principles of percutaneous absorption, New
 1202 York, Karger.

1203

1204 Shah U.U., Roberts M., Gul M.O., Teleu C., Beresford M.W., (2011). Needle-free and
 1205 microneedle drug delivery in children: a case for disease-modifying antirheumatic drugs
 1206 (DMARDs), *International Journal of Pharmaceutics*, **416**: 1-11

1207

1208 Smith F.D., Harpending P.R., Sanford J.C., (1992). Biolistic transformation of prokaryotes:
 1209 factors that affect biolistic transformation of very small cells. *J. Gen. Microbiol.*, **138**:239-248

1210

1211 Silpi C., Manish B., Kumar T.R., (2011). Microneedles in transdermal drug delivery: An unique
 1212 painless option, *International Research Journal of Pharmacy*, **2**(4): 72-78

1213

1214 Singh R., Dahotre N.B., (2007). Corrosion degradation and prevention by surface modification
 1215 of biometallic materials, *J. Mater. Sci. Mater. Med.*, **18**(5):725-751

1216

1217 Soliman S.M. (2011). Micro-particle and Gas Dynamics in an Axi-symmetric Supersonic
1218 Nozzle, University of Cincinnati (Cincinnati, USA), *Thesis for the degree of Doctor of*
1219 *Philosophy in Aerospace Engineering*.
1220
1221 Soliman S.M., Abdallah S., (2011a). CFD investigation of powdered vaccine and gas dynamics
1222 in biolistic gun, *Powder Technology*, **214**: 135-142
1223
1224 Soliman S.M., Abdallah S., Gutmark E., Turner M.G., (2011b). Numerical simulation of
1225 microparticles penetration and gas dynamics in an axi-symmetric supersonic nozzle for
1226 genetic vaccination, *Powder Technology*, **208**: 676-683
1227
1228 Stoeber B., Liepmann D., (2000). Fluid injection through out-of-plane microneedles,
1229 *Proceedings of the 1st Annual International IEEE-EMBS Special Topic Conference on*
1230 *Microtechnologies in Medicine and Biology, Lyon, France*, **14**: 224-228.
1231
1232 Sung H.W., Chen M.C., Lee P.W., Tu H.S., (2011). Nanoparticles for protein/peptide delivery
1233 and delivery means thereof, Patent No. US7901711 B1 (Granted), United States Patent, USA
1234
1235 Svarovsky S., Borovkov A., Sykes K., (2008). Cationic gold microparticles for biolistic delivery
1236 of nucleic acids, *Biotechniques*, **45**(5): 535-540
1237
1238 Tekeuchi Y., Doston M., Keen N.T., (1992). Plant transformation: a simple particle
1239 bombardment device based on flowing helium, *Plant Mol. Biol.*, **18**: 835-839
1240
1241 Thomas J.L., Bardou J., Mauchamp B. et al, Chavancy Gerard (2001). A helium burst biolistic
1242 device adapted to penetrate fragile insect tissues, *Journal of Insect Science*, **1**(9): 1-10
1243
1244 Tortora G.J., Derrickson B.H., (2011). Principles of Anatomy and Physiology, 13 edition, Wiley,
1245 ISBN-13: 978-0470565100
1246

1247 Trainer A.T., Alexander M.Y., (1997). Gene delivery to the epidermis, *Human Molecular*
 1248 *Genetics*, **6**(10): 1761 - 1767
 1249
 1250 Truong N.K., Liu Y., Kendall M.A.F., (2006). Gas and particle dynamics of a contoured shock
 1251 tube for pre-clinical micro-particle drug delivery, *Shock Waves*, **15**: 149-164
 1252
 1253 Tuan-Mahmood T.M., McCrudden M.T.C., Torrisi B.M., McAlister E., Garland M.J., Singh
 1254 T.R.R., Donnelly R.F., (2013). Microneedles for intradermal and transdermal drug delivery,
 1255 *European Journal of Pharmaceutical Sciences*, doi: 10.1016/j.ejps.2013.05.005. (in press)
 1256
 1257 Uchida M., Li X.W., Mertens P., Alpar H.O., (2009). Transfection by particle bombardment:
 1258 Delivery of plasmid DNA into mammalian cells using gene gun, *Biochim. Biophys. Acta.*,
 1259 **1790**(8): 754-764
 1260
 1261 Valenstein J.C., (2012). Developing nanotechnology for biofuel and plant science applications,
 1262 Iowa State Univeristy (USA), *Thesis for the degree of Doctor of Philosophy*
 1263
 1264 Verbaan, F.J., Bal, S.M., van den Berg, D.J., Dijkman, J.A., van Hecke, M., Verpoorten, H.,
 1265 van den Berg, A., Luttge, R., Bouwstra, J.A. (2008). Improved piercing of microneedle arrays
 1266 in dermatomed human skin by an impact insertion method, *Journ. Contrl. Rel.*, **128**: 80–88.
 1267
 1268 Williams R.S., Johnston S.A., Riedy M., DeVit M.J., McElligott S.G., Sanford J.C., (1991).
 1269 Introduction of foreign genes into tissues of living mice by DNA-coated microprojectiles, *Proc.*
 1270 *Natl. Acad. Sci. U.S.A.*, **88**: 2726-2730
 1271
 1272 Xia J.X., Martinez A., Daniell H., Ebert S.N., (2011). Evaluation of biolistic gene transfer
 1273 methods *in vivo* using non-invasive bioluminescent imaging techniques, *BMC Biotechnology*,
 1274 **11**:62
 1275

1276 Yager E.J., Stagnar C., Gopalakrishnan R., Fuller J.T., Fuller D.H., (2013). Optimizing
 1277 particle-mediated epidermal delivery of an influenza DNA vaccine in ferrets, *Methods Mol. Biol.*,
 1278 **940**: 223-237
 1279
 1280 Yang N.S., Burkholder J., Roberts B., Martinell B., McCabe D., (1990). *In vivo* and *in vitro* gene
 1281 transfer to mammalian somatic cells by particle bombardment, *Proc. Natl. Acad. Sci. U.S.A.*,
 1282 **87**: 9568-9572
 1283
 1284 Yen M.C., Lai M.D., (2013). Biolistic DNA delivery to mice with the low pressure gene gun,
 1285 *Methods Mol. Biol.*, **940**:169-174
 1286
 1287 Yoshida Y., Kobayashi E., Endo H., Hamamoto T., Yamanaka T., Fujimura A., Kagawa Y.,
 1288 (1997). Introduction of DNA into rat liver with a hand-held gene gun: Distribution of the
 1289 expressed enzyme, [32P] DNA, and Ca²⁺ flux, *Biochem. Biophys. Res. Commun.* **234**(3): 695
 1290 -700.
 1291
 1292 Yoshimitsu Y., Tanaka K., Tagawa T., Nakamura Y., Matsuo T., Okamoto S., (2009),
 1293 Improvement of DNA/Metal Particle Adsorption in Tungsten-Based Biolistic Bombardment;
 1294 Alkaline pH is Necessary for DNA Adsorption and Suppression of DNA Degradation, *Journal*
 1295 *of Plant Biology*, **52**: 524-532.
 1296
 1297 Yuzhakov V.V., (2008). Tissue conforming microneedle array and patch for transdermal drug
 1298 delivery or biological fluid collection, *World Intellectual Property Organization*, Patent: WO
 1299 2008/067290 A2;
 1300
 1301 Zhang M.J., Tao W.M., Pianetta P.A., (2007). Dynamics modeling of biolistic gene guns, *Phys.*
 1302 *Med. Biol.*, **52**: 1485-1493
 1303
 1304 Zhang D.W., Das D.B., Rielly C.D., (2013a). An experimental study of microneedles assisted
 1305 micro-particle delivery using a model system, *Journal of Pharmaceutical Science*, **102**(10):
 1306 3632-3644, DOI: 10.1002/jps.23665

1307

1308 Zhao H.Y., Avenarius M.R., Gillespie P.G., (2012). Improved biolistic transfection of hair cells,
 1309 *PLoS ONE*, **7**(10): e46765

1310

1311 Zhao Y., J., Zhang J.,Y., Yang X.,L., (2006). "Design and Manufacture of Microneedles Array
 1312 for Transdermal Drug Delivery", *Chinese Journal of Medical Instrumentation*, **30**(1): 33-38

1313

1314 Zhou F., (2000). High-pressure Gas Gene Gun, Chinese Patent No. CN 1262323A (Granted),
 1315 *State Intellectual Property Office of the P.R. China*, China

1316

1317 Zhou F., (2007). Liquid Gene Gun, Chinese Patent No.CN 189217A (Granted), *State*
 1318 *Intellectual Property Office of the P.R. China*, China

1319

1320 Zhou W.Z., (1995). Gene Gun, Chinese Patent No. CN 2196119Y (Granted), *State Intellectual*
 1321 *Property Office of the P.R. China*, China

1322

1323 Zhu, M., Chen, Y., Wang, Z. Lu, Y., Ge, H., Tang, Y. (2012). Fabrication Method for Hollow
 1324 Microneedles for Drug Delivery. Nanjing University, Patent US 8137736 B2 (Granted), *United*
 1325 *States Patent*, USA

1326

1327 Ziegler A.S., (2008). Needle-free delivery of powdered protein vaccine: a new and rapidly
 1328 developing technique, *J. Pharm. Innov.*, **3**: 204-213

1329

1330 Zilony N., Tzur-Balter A., Segal E., Shefi O., (2013). Bombarding Cancer: Biolistic delivery of
 1331 therapeutics using porous Si carriers, *Scientific Reports*, **3**: article number 2499

1332

1333 Zuraida A.R., Rahiniza K., Hafiza M.R.N., Roowi S., Zamri Z., Subramaniam S., (2010).
 1334 Factors affecting delivery and transient expression of gusa gene in Malaysian indica rice MR
 1335 219 callus via biolistic gun system, *African Journal of Biotechnology*, **9**(51):8819-8818.